

Consult Instruction for Use.

APPLICATION COBAS MIRA - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

GENERAL	
Measurement Mode	: ABSORB
Reaction Mode	: R-S
Calibration Mode	: CALIBRATOR
Reagent Blank	: REAG/DIL
Cleaner	: NO
Wavelength	: 600 nm
Decimal position	: 2
Unit	: g/dL
ANALYSIS	
Sample Dil..Name	: H ₂ O
Post Dil.Factor	: NO
Conc.Factor	: NO
Sample	Cycle 1
Vol: 2.0 μL	Dil: 10 μL
Reagent	Cycle 1
Vol: 300 μL	
CALCULATION	
Sample limit	: NO
Reac.Direction	: INCREASE
Check	: ON
Convers.Factor	: 1.00000
Offset	: 0.00000
Test Range Low	: *
High	: *
Norm.Range Low	: *
High	: *
Number of steps	: 1
Calc.Step A	: ENDPOINT
Readings First	: CB
Last	: 2
Reaction Limit	: NO
CALIBRATION	
Calib.Interval	: ON REQUEST
BLANK	
Reag.Range Low	: NO
High	: NO
Blank Range Low	: NO
High	: NO
STANDARD POS	: As desired
VAL	: **
REPL	: TRIPL
DEV	: 5%
* :	Entered by the operator.
** :	Entered calibrator value.