

Consult Instruction for Use.

APPLICATION COBAS MIRA - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

GENERAL

Measurement Mode : ABSORB
 Reaction Mode : R-S
 Calibration Mode : CALIBRATOR
 Reagent Blank : REAG/DIL
 Cleaner : NO
 Wavelength : 500 nm
 Decimal position : 0
 Unit : mg/dL

ANALYSIS

Sample Dil.Name : H₂O
 Post Dil.Factor : NO
 Conc.Factor : NO
Sample
 Cycle 1
 Vol : 3 µL Dil : 10 µL
Reagent
 Cycle 1
 Vol : 290 µL

CALCULATION

Sample limit : NO
 Reac.Direction : INCREASE
 Check : ON
 Convers.Factor : 1.00000
 Offset : 0.00000
 Test Range Low : *
 High : *
 Norm.Range Low : *
 High : *
 Number of steps : 1
 Calc.Step A : ENDPOINT
 Readings First : CB
 Last : 17
 Reaction Limit : NO

CALIBRATION

Calib.Interval : ON REQUEST
BLANK
 Reag.Range Low : NO
 High : NO
 Blank Range Low : NO
 High : NO
STANDARD POS
 : As desired
 VAL : **
 REPL : TRIPL
 DEV : 5%

* : Entered by the operator.

** : Entered calibrator value