

HITACHI 911

AA-H911-CAT-11

BIOCHEMISTRY

A		Versions	I		Versions		
	ACID PHOSPHATASE	PACI-1	PRO	IRON FERROZINE	FEFR -4	PRO	
	ALBUMIN	ALBU-4	PRO				
	ALP (DEA) SL mono procedure	PASLmono-4	PRO				
	ALP (DEA) SL bi procedure	PASLbi-4	PRO				
	ALP (DEA)	PALC-3	PRO				
	ALT / GPT SL 4+1 mono procedure ...	ALSL4+1mono-2 ..	PRO				
	ALT / GPT SL 4+1 bi procedure	ALSL4+1bi-2	PRO	<i>Update</i>	LDH - P SL 4+1 mono procedure	LDSL4+1mono-2 PRO	
<i>Update</i>	ALT / GPT	ALAT-4	PRO		LDH - P SL 4+1 bi procedure	LDSL4+1bi-2	PRO
	AMYLASE SL	AMSL-4	PRO		LDH - P	LDHP-4	PRO
	AST / GOT SL 4+1 mono procedure...	ASSL4+1mono-2...	PRO				
	AST / GOT SL 4+1 bi procedure	ASSL4+1bi-2	PRO				
<i>Update</i>	AST / GOT	ASAT-4	PRO				
B			P				
	BILIRUBIN TOTAL & DIRECT 4+1	BITD-1	PRO	PHOSPHORUS	PHOS-4	PRO	
C			T				
	CALCIUM ARSENAZO	CALA-4	PRO				
	CALCIUM OCPC mono procedure	CALOmono-3	PRO				
	CHLORIDE	CHLO-4	PRO	<i>Update</i>	TOTAL PROTEIN	PRTB-4	PRO
<i>Update</i>	CHOLESTEROL SL	CHSL 5	PRO	<i>Update</i>	TRIGLYCERIDES MONO SL NEW ...	TGMLN-3	PRO
<i>Update</i>	CHOLESTEROL	CHOL-4	PRO		TRIGLYCERIDES	TRIG-4	PRO
<i>Update</i>	CHOLESTEROL HDL	HDLC-5	PRO				
<i>NEW</i>	CHOLESTEROL HDL SL 2G	HDLL-1	PRO				
<i>NEW</i>	CHOLESTEROL LDL SL 2G	LDLL-1	PRO				
<i>Update</i>	CHOLINESTERASE	CHES-5	PRO				
	CK NAC SL mono procedure	CKSLmono-5	PRO				
<i>Update</i>	CK NAC	CKNA-4	PRO	<i>Update</i>	UREA UV SL mono procedure	URSLmono-5	PRO
	CK - MB SL mono procedure	CMSLmono-4	PRO	<i>Update</i>	UREA UV SL bi procedure	URSLbi-5	PRO
	CK - MB	CKMB-3	PRO		UREA UV	URUV-3	PRO
	COPPER mono procedure	CUIVmono-4	PRO	<i>Update</i>	URIC ACID MONO SL	AUML-2	PRO
	COPPER bi procedure	CUIVbi-3	PRO		URIC ACID SL mono procedure	AUSLmono-3	PRO
	CREATININE JAFFE mono procedure	CRCOmono-4	PRO		URIC ACID SL bi procedure	AUSLbi-3	PRO
<i>NEW</i>	CREATININE PAP SL	CRSL-1	PRO	<i>Update</i>	URIC ACID	ACUR-4	PRO

G

	GAMMA GT SL mono procedure	GASLmono-4	PRO
	GAMMA GT SL bi procedure	GASLbi-4	PRO
	GAMMA GT	GAGT-3	PRO
	GLUCOSE HK SL mono procedure	GHSLmono-3	PRO
	GLUCOSE HK SL bi procedure	GHSLbi-3	PRO
<i>Update</i>	GLUCOSE PAP SL	GPSL-5	PRO
	GLUCOSE PAP	GLUP-3	PRO

SPECIFIC PROTEINS

<i>NEW</i>	APO A1 IP	IAPA-3	PRO
<i>NEW</i>	APO B IP	IAPB-3	PRO
<i>NEW</i>	CRP IP	ICRP-1	PRO
<i>NEW</i>	HAPTOGLOBIN IP	IHAP-1	PRO
<i>NEW</i>	HbA1c	HBAC-1	PRO
<i>NEW</i>	HbA1c on board lyse	HBACobl-1	PRO
<i>NEW</i>	µALBUMIN IP	IMAL-1	CE
<i>NEW</i>	IgA IP	IIGA-1	CE
<i>NEW</i>	IgG IP	IIGG-1	CE
<i>NEW</i>	IgM IP	IIGM-1	CE
<i>NEW</i>	OROSOMUCOID IP	IORO-1	PRO
<i>NEW</i>	PREALBUMIN IP	IPAL-1	PRO
<i>NEW</i>	TRANSFERRIN IP	ITRF-1	PRO

PRO = PROPOSAL: These applications have been established theoretically. It is advised to check it.



ACID PHOSPHATASE

Ref. : PACI-0030

18 x 3 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use : 

Working temperature : 37°C

TEST	(ACP) (*)	TEST NAME	(ACP)	UNITS	(UI/L)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(ACID PHOSPHATASE)						
CONTROL INTERVAL	(1000)	INSTRUM. FACTOR	(Y=aX+b) a (1.0) b (0)						
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	AGE M F								
	(0) (Y) (0) (0) (0) (0)		(0) (0)						
	(0) (Y) (0) (0) (0) (0)								
	(\$\$) (\$\$)		(\$\$) (\$\$)						
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(40)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(20)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ACP)	ASSAY CODE (RATE A)(10)(-)				WAVELENGTH (SUB/MAIN) (---)/(415)				
	ASSAY POINT (19) (31) (0) (0)				DILUTION (00311) (0)				
S. VOLUME (REGULAR)	< S.TYPE 1 >				< S.TYPE 2 >				
	(20) (0) (0)				(1) (0) (0)				
S. VOLUME (DECREASE)	#) (0) (0)				(1) (0) (0)				
S. VOLUME (INCREASE)	#) (0) (0)				(1) (0) (0)				
ABS. LIMIT	(-3200) (INCREASE) ()								
PROZONE LIMIT	(32000) (LOWER)								
REAGENT	R1	(250) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE	(LINEAR) (1) (0) (0)								
AUTO CALIB.	Please note: Set K factor to: [1553] Adjust if necessary								
TIME OUT BLANK	(0)	STD LIMIT	(0.1)						
SPAN	(0)	DUPLICATE LIMITE	(1000)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(-32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter position or code number. *** : Enter calibrator or standard value.
 \$\$\$: Enter values.

: Denotes a user-defined parameter. Information is specific to the channel being used or, it is determined by the laboratory's own preference for operation.



ALBUMIN

Ref. : ALBU-0600 2 x 125 mL
 Ref. : ALBU-0700 4 x 250 mL

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (ALB) (*)		TEST NAME (ALB)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (ALBUMIN)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)		(0) (0) (0) (0)		(0) (0)					
(0) (Y) (0) (0) (0) (0)		(0) (0) (0) (0)							
(\$\$) (\$\$) (\$\$) (\$\$)		(\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(69)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALB)		WAVELENGTH (SUB/MAIN) (700)/(600)							
ASSAY CODE (1 POINT)(10)()		DILUTION (00301) (99)							
ASSAY POINT (31) (0) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(2) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(6) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(0) () (LOWER)							
REAGENT									
R1		(300) (0) (*) (0)							
R2		(0) (0) (*) (0)							
R3		(0) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT (0.1)							
SPAN (0)		DUPLICATE LIMITE (700)							
2 POINT (0)		SENSITIVITY LIMIT (0)							
FULL (0)		S1 ABS LIMIT (- 32000) (32000)							
CHANGE LOT (CANCEL)		COMPENSATED LIMIT ()							
BOTTLE (CANCEL)									

* : Enter position or code number. *** : Enter calibrator or standard value.
 \$\$\$: Enter expected values.



ALP (DEA) SL

ONE REAGENT PROCEDURE

Ref. : PASL-0400 2 x 62.5 mL
 PASL-0420 4 x 62.5 mL
 PASL-0500 5 x 125 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (ALP) (*)		TEST NAME (ALP)		UNITS (UI/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (ALKALINE PHOSPHATASE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(800)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(5)	(0)	(0)	(*)	(1)	(0)	()
(2)	(**)	(*)	(5)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALP)		WAVELENGTH (SUB/MAIN) (700)/(415)							
ASSAY CODE (KIN A)(10)()		DILUTION (00301) (99)							
ASSAY POINT (7) (17) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(5) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(10) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(11000) () (2 : INCREASE)							
PROZONE LIMIT		(0) () (1 : HIGHER)							
REAGENT		(250) (0) (*) (0)							
	R1	(0) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter values.



ALP (DEA) SL

Ref. : PASL-0400 2 x 62.5 mL
 PASL-0420 4 x 62.5 mL
 PASL-0500 5 x 125 mL

TWO REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (ALP) (*)		TEST NAME (ALP)		UNITS (UI/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (ALKALINE PHOSPHATASE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)			(0) (0)				
(0) (Y)	(0) (0)	(0) (0)							
	(\$\$) (\$\$)	(\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(800)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(4)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(4)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALP)		ASSAY CODE (KIN A)(10)()		WAVELENGTH (SUB/MAIN) (700)/(415)					
ASSAY POINT (19) (31) (0) (0)				DILUTION (00301) (99)					
S. VOLUME (REGULAR)			< S.TYPE 1 >		< S.TYPE 2 >				
			(4) (0) (0)		(1) (0) (0)				
S. VOLUME (DECREASE)			(1) (0) (0)		(1) (0) (0)				
S. VOLUME (INCREASE)			(8) (0) (0)		(1) (0) (0)				
ABS. LIMIT			(11000) () (2 : INCREASE)						
PROZONE LIMIT			(0) () (1 : HIGHER)						
REAGENT		R1	(200) (0) (*) (0)						
		R2	(0) (0) (*) (0)						
		R3	(50) (0) (*) (0)						
		R4	(0) (0) (*) (0)						
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter values.



ALP (DEA)

Ref. : PALC-0030 20 x 3 mL
 PALC-0200 12 x 20 mL

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (ALP) (*)		TEST NAME (ALP)		UNITS (UI/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (ALKALINE PHOSPHATASE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(800)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(5)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(5)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALP)									
ASSAY CODE (KIN A)(10)()			WAVELENGTH (SUB/MAIN) (700)/(415)						
ASSAY POINT (7) (17) (0) (0)			DILUTION (00301) (99)						
			< S.TYPE 1 >		< S.TYPE 2 >				
S. VOLUME (REGULAR)			(5) (0) (0)		(1) (0) (0)				
S. VOLUME (DECREASE)			(3) (0) (0)		(1) (0) (0)				
S. VOLUME (INCREASE)			(10) (0) (0)		(1) (0) (0)				
ABS. LIMIT			(11000) () (2 : INCREASE)		(1) (0) (0)				
PROZONE LIMIT			(0) () (1 : HIGHER)		(1) (0) (0)				
REAGENT R1			(250) (0) (*) (0)		(1) (0) (0)				
R2			(0) (0) (*) (0)		(1) (0) (0)				
R3			(0) (0) (*) (0)		(1) (0) (0)				
R4			(0) (0) (*) (0)		(1) (0) (0)				
CALIBRATION TYPE						(LINEAR) (2) (2) (0) ()			
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT		(0.1)				
SPAN (0)			DUPLICATE LIMITE		(200)				
2 POINT (0)			SENSITIVITY LIMIT		(0)				
FULL (0)			S1 ABS LIMIT		(-32000) (32000)				
CHANGE LOT (CANCEL)			COMPENSATED LIMIT		()				
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter values.



ALT / GPT 4 + 1 SL

Ref. : ALSL-0410 2 x 62.5 mL
 ALSL-0430 4 x 62.5 mL
 ALSL-0510 5 x 125 mL

ONE REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (ALAT) (*)		TEST NAME (ALAT)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (ALAT)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)				(0) (0)					
(0) (Y) (0) (0) (0) (0)									
		(\$\$) (\$\$) (\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(300)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALAT)						WAVELENGTH (SUB/MAIN) (700)/(340)			
ASSAY CODE (KIN A)(10)()						DILUTION (00301) (99)			
ASSAY POINT (7) (17) (0) (0)									
S. VOLUME (REGULAR)						< S.TYPE 1 >		< S.TYPE 2 >	
						(20) (0) (0)		(1) (0) (0)	
S. VOLUME (DECREASE)						(3) (0) (0)		(1) (0) (0)	
S. VOLUME (INCREASE)						(40) (0) (0)		(1) (0) (0)	
ABS. LIMIT						(4000) () (1 : DECREASE)			
PROZONE LIMIT						(0) () (2 : LOWER)			
REAGENT R1						(200) (0) (*) (0)			
R2						(0) (0) (*) (0)			
R3						(0) (0) (*) (0)			
R4						(0) (0) (*) (0)			
CALIBRATION TYPE						(LINEAR) (2) (2) (0) ()			
AUTO CALIB.									
TIME OUT BLANK (0)						STD LIMIT		(0.1)	
SPAN (0)						DUPLICATE LIMITE		(200)	
2 POINT (0)						SENSITIVITY LIMIT		(0)	
FULL (0)						S1 ABS LIMIT		(-32000) (32000)	
CHANGE LOT (CANCEL)						COMPENSATED LIMIT		()	
BOTTLE (CANCEL)									

* : Enter position or code number
 \$\$\$: Enter values

*** : Enter calibrator or standard value



ALT / GPT 4 + 1 SL

Ref. : ALSL-0410 2 x 62.5 mL
 : ALSL-0430 4 x 62.5 mL
 : ALSL-0510 5 x 125 mL

TWO REAGENT PROCEDURE

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use : **i**

Working temperature : 37°C

TEST (ALAT) (*)		TEST NAME (ALAT)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (ALAT SL TWO)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
(0)-(300)			(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALAT)									
ASSAY CODE (KIN A)(10)()		WAVELENGTH (SUB/MAIN) (700)/(340)							
ASSAY POINT (19) (31) (0) (0)		DILUTION (00301) (99)							
				< S.TYPE 1 >		< S.TYPE 2 >			
S. VOLUME (REGULAR)				(25) (0) (0)		(1) (0) (0)			
S. VOLUME (DECREASE)				(3) (0) (0)		(1) (0) (0)			
S. VOLUME (INCREASE)				(40) (0) (0)		(1) (0) (0)			
ABS. LIMIT				(4000) () (1 : DECREASE)					
PROZONE LIMIT				(0) () (2 : LOWER)					
REAGENT		R1	(200) (0) (*) (0)						
		R2	(0) (0) (*) (0)						
		R3	(50) (0) (*) (0)						
		R4	(0) (0) (*) (0)						
CALIBRATION TYPE				(LINEAR) (2) (2) (0) ()					
AUTO CALIB.									
TIME OUT BLANK		(0)	STD LIMIT		(0.1)				
SPAN		(0)	DUPLICATE LIMITE		(200)				
2 POINT		(0)	SENSITIVITY LIMIT		(0)				
FULL		(0)	S1 ABS LIMIT		(-32000) (32000)				
CHANGE LOT		(CANCEL)	COMPENSATED LIMIT		()				
BOTTLE		(CANCEL)							

* : Enter position or code number
 \$\$\$: Enter expected values.

*** : Enter calibrator or standard value

ALT / GPT

Ref.: ALAT-0200 12 x 20 mL
ALAT-0400 9 x 50 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use : 


Working temperature : 37°C

TEST	(ALAT) (*)	TEST NAME	(ALAT)	UNITS	(U/L)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(ALAT)						
CONTROL INTERVAL	(300)	INSTRUM. FACTOR (Y=aX+b)	a (1.0) b (0)						
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(300)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALAT)		WAVELENGTH (SUB/MAIN)	(700)/(340)						
ASSAY CODE (KIN A)	(10) ()	DILUTION (00301)	(99)						
ASSAY POINT (7) (17) (0) (0)									
S. VOLUME (REGULAR)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (DECREASE)		(20) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(3) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(40) (0) (0)		(1) (0) (0)					
PROZONE LIMIT		(4000) () (1 : DECREASE)							
REAGENT	R1	(0) () (2 : LOWER)							
	R2	(250) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(0.1)						
SPAN	(0)	DUPLICATE LIMITE	(200)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(-32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter position or code number

*** : Enter calibrator or standard value

§§§ : Enter expected values.

 Modification from the previous version

(06/2009)
AA-H911-ALAT-4



AMYLASE SL

Ref. : AMSL-0390 1 x 50 mL
 AMSL-0400 6 x 50 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (AMY) (*)		TEST NAME (AMY)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (AMYLASE SL)							
CONTROL INTERVAL (0)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)				(0) (0)					
(0) (Y) (0) (0) (0) (0)									
		(§§) (§§) (§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(1200)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(4)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(4)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (AMY)		WAVELENGTH (SUB/MAIN) (660)/(415)							
ASSAY CODE (6 KINETIC A)(10)()		DILUTION (00301) (99)							
ASSAY POINT (6) (17) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(4) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(1) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(8) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(8000) () (2 : INCREASE)							
PROZONE LIMIT		(0) () (1 : HIGHER)							
REAGENT R1		(250) (0) (*) (*)							
R2		(0) (0) (*) (*)							
R3		(0) (0) (*) (*)							
R4		(0) (0) (*) (*)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(- 32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.
 §§§ : Enter expected values.

*** : Enter calibrator or standard value.



AST / GOT 4 + 1 SL

Ref. : ASSL-0410 2 x 62.5 mL
 ASSL-0430 4 x 62.5 mL
 ASSL-0510 5 x 125 mL

ONE REAGENT PROCEDURE

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (ASAT) (*)		TEST NAME (ASAT)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (ASAT SL ONE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)				(0) (0)					
(0) (Y) (0) (0) (0) (0)									
(\$\$) (\$\$) (\$\$) (\$\$)									
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(300)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ASAT)		WAVELENGTH (SUB/MAIN) (700)/(340)							
ASSAY CODE (KIN A) (10)()		DILUTION (00301) (99)							
ASSAY POINT (7) (17) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(20) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(40) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(4000) () (1 : DECREASE)							
PROZONE LIMIT		(0) () (2 : LOWER)							
REAGENT R1		(200) (0) (*) (0)							
R2		(0) (0) (*) (0)							
R3		(0) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number
 \$\$\$: Enter values

*** : Enter calibrator or standard value



AST / GOT 4 + 1 SL

Ref. : ASSL-0410 2 x 62.5 mL
 ASSL-0430 4 x 62.5 mL
 ASSL-0510 5 x 125 mL

TWO REAGENT PROCEDURE

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (ASAT) (*)		TEST NAME (ASAT)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (ASAT SL TWO)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)				(0) (0)					
(0) (Y) (0) (0) (0) (0)									
		(\$\$) (\$\$) (\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(300)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ASAT)		WAVELENGTH (SUB/MAIN) (700)/(340)							
ASSAY CODE (KIN A) (10)()		DILUTION (00301) (99)							
ASSAY POINT (19) (31) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(25) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(40) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(4000) () (1 : DECREASE)							
PROZONE LIMIT		(0) () (2 : LOWER)							
REAGENT R1		(200) (0) (*) (0)							
R2		(0) (0) (*) (0)							
R3		(50) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number
 \$\$\$: Enter values

*** : Enter calibrator or standard value



AST / GOT

☞ Ref.: ASAT-0200 12 x 20 mL
ASAT-0400 9 x 50 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (ASAT) (*)		TEST NAME (ASAT)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (ASAT)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(300)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ASAT)			WAVELENGTH (SUB/MAIN) (700)/(340)						
ASSAY CODE (KIN A)(10)()			DILUTION (00301) (99)						
ASSAY POINT (7) (17) (0) (0)									
S. VOLUME (REGULAR)			< S.TYPE 1 >	< S.TYPE 2 >					
			(20) (0) (0)	(1) (0) (0)					
S. VOLUME (DECREASE)			(3) (0) (0)	(1) (0) (0)					
S. VOLUME (INCREASE)			(40) (0) (0)	(1) (0) (0)					
ABS. LIMIT			(4000) () (1 : DECREASE)						
PROZONE LIMIT			(0) () (2 : LOWER)						
REAGENT		R1	(250) (0) (*) (0)						
		R2	(0) (0) (*) (0)						
		R3	(0) (0) (*) (0)						
		R4	(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK		(0)	STD LIMIT		(0.1)				
SPAN		(0)	DUPLICATE LIMITE		(200)				
2 POINT		(0)	SENSITIVITY LIMIT		(0)				
FULL		(0)	S1 ABS LIMIT		(- 32000) (32000)				
CHANGE LOT		(CANCEL)	COMPENSATED LIMIT		()				
BOTTLE		(CANCEL)							

* : Enter position or code number.
§§§ : Enter values.

*** : Enter calibrator or standard value.

Modification from the previous version

(06/2009)
AA-H911-ASAT-4




BILIRUBIN

TOTAL & DIRECT

4+1

Ref.: BITD-0600 TOTAL & DIRECT 4+1 2 x 125 mL
 BIDI-0600 DIRECT 4+1 2 x 125 mL
 BITO-0600 TOTAL 4+1 2 x 125 mL

APPLICATION HITACHI 911
PROPOSAL

Instruction use : 
Working temperature : 37°C
Preparation : Reagent 1: Ready for use
 Reagent 2: Ready for use

TEST (BITD) (*)		TEST NAME (BITD)			UNITS (mg/dL)				
DATA MODE (1: ON BOARD)		REPORT NAME (BILIRUBIN)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE			<S.TYPE 2>				
AGE M F									
(0) (Y) (0) (0) (0) (0)					(0) (0)				
(0) (Y) (0) (0) (0) (0)									
		(§§) (§§) (§§) (§§)							
TECHNICAL LIMIT		EXPECTED VALUE			<S.TYPE 2>				
					(0)-(0)				
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(30)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(30)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (BITD)									
ASSAY CODE (2:2POINT-END)(10)()					WAVELENGTH (SUB/MAIN) (no)/(546)				
ASSAY POINT (15) (31) (0) (0)					DILUTION (00301) (99)				
					< S.TYPE 1 >		< S.TYPE 2 >		
S. VOLUME (REGULAR)					(30) (0) (0)		(1) (0) (0)		
S. VOLUME (DECREASE)					#) (0) (0)		(1) (0) (0)		
S. VOLUME (INCREASE)					#) (0) (0)		(1) (0) (0)		
ABS. LIMIT					(0) () (2 : INCREASE)				
PROZONE LIMIT					(-32000) () (2 : LOWER)				
REAGENT R1					(240) (0) (*) (0)				
R2					(0) (0) (*) (0)				
R3					(60) (0) (*) (0)				
R4					(0) (0) (*) (0)				
CALIBRATION TYPE					(LINEAR) (2) (2) (0) ()				
AUTO CALIB.									
TIME OUT BLANK (0)					STD LIMIT		(999.9)		
SPAN (0)					DUPLICATE LIMITE		(500)		
2 POINT (0)					SENSITIVITY LIMIT		(0)		
FULL (0)					S1 ABS LIMIT		(-32000) (32000)		
BOTTLE (CANCEL)									

*: Enter position or code number *** : Enter calibrator or standard value
 §§: Enter expected values
 #: User defined



CALCIUM ARSENAZO

Ref. : CALA-0600 2 x 125 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use : 

Working temperature : 37°C

TEST	(Ca) (*)	TEST NAME	(Ca)	UNITS	(mg/l)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(CALCIUM ARSENAZO)						
CONTROL INTERVAL	(80)	INSTRUM. FACTOR	(Y=aX+b) a (1.0) b (0)						
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	AGE M F								
	(0) (Y) (0) (0) (0) (0)		(0) (0)						
	(0) (Y) (0) (0) (0) (0)								
	(\$\$) (\$\$) (\$\$) (\$\$)								
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(150)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(**)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (Ca)									
ASSAY CODE	(1 POINT)(10)()			WAVELENGTH (SUB/MAIN)	(700)/(660)				
ASSAY POINT	(31) (0) (0) (0)			DILUTION	(00301) (99)				
S. VOLUME (REGULAR)	< S.TYPE 1 >			< S.TYPE 2 >					
	(3) (0) (0)			(1) (0) (0)					
S. VOLUME (DECREASE)	(1) (0) (0)			(1) (0) (0)					
S. VOLUME (INCREASE)	(6) (0) (0)			(1) (0) (0)					
ABS. LIMIT	(0) () (INCREASE)								
PROZONE LIMIT	(32000) () (UPPER)								
REAGENT	R1	(300) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE	(LINEAR) (2) (2) (0) ()								
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(0.1)						
SPAN	(0)	DUPLICATE LIMITE	(1000)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(- 32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



CALCIUM OCPC

Ref. : CALO-0600 2 x 125 mL

ONE REAGENT PROCEDURE

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (Ca) (*)		TEST NAME (Ca)		UNITS (mg/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (CALCIUM OCP)							
CONTROL INTERVAL (80)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(135)							
					(0)-(0)				
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (Ca)									
ASSAY CODE (1 POINT)(10)()			WAVELENGTH (SUB/MAIN) (660)/(570)						
ASSAY POINT (31) (0) (0) (0)			DILUTION (00301) (99)						
			< S.TYPE 1 >			< S.TYPE 2 >			
S. VOLUME (REGULAR)			(3) (0) (0)			(1) (0) (0)			
S. VOLUME (DECREASE)			(1) (0) (0)			(1) (0) (0)			
S. VOLUME (INCREASE)			(6) (0) (0)			(1) (0) (0)			
ABS. LIMIT			(0) () (INCREASE)						
PROZONE LIMIT			(32000) () (UPPER)						
REAGENT			(300) (0) (*) (0)						
R1			(0) (0) (*) (0)						
R2			(0) (0) (*) (0)						
R3			(0) (0) (*) (0)						
R4			(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT			(0.1)			
SPAN (0)			DUPLICATE LIMITE			(1000)			
2 POINT (0)			SENSITIVITY LIMIT			(0)			
FULL (0)			S1 ABS LIMIT			(- 32000) (32000)			
CHANGE LOT (CANCEL)			COMPENSATED LIMIT			()			
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



CHLORIDE

Ref. : CHLO-0600 2 x 125 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (CHL) (*)		TEST NAME (CHL)		UNITS (mEq/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (CHLORIDES)							
CONTROL INTERVAL (80)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)	(0) (0)					
	(0) (Y)	(0) (0)	(0) (0)						
		(§§) (§§)	(§§) (§§)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(130)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(5)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(5)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CHL)									
ASSAY CODE (1 POINT)(10)()			WAVELENGTH (SUB/MAIN) (700)/(480)						
ASSAY POINT (31) (0) (0) (0)			DILUTION (00301) (99)						
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(5) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(10) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(32000) () (UPPER)							
REAGENT		(300) (0) (*) (0)							
	R1	(0) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(700)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



CHOLESTEROL SL

Ref. : CHSL-0490	1 x 100 mL
CHSL-0500	6 x 100 mL
CHSL-0700	4 x 250 mL
☞ CHSL-0507	6 x 100 mL + STD
☞ CHSL-0707	4 x 250 mL + STD

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (CHOL) (*)		TEST NAME (CHOL)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (CHOLESTEROL SL)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)		(0) (0)					
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(6)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CHOL)									
ASSAY CODE (1 POINT)(10) ()			WAVELENGTH (SUB/MAIN) (700)/(505)						
ASSAY POINT (31) (0) (0) (0)			DILUTION (00301) (99)						
			< S.TYPE 1 >			< S.TYPE 2 >			
S. VOLUME (REGULAR)			(3) (0) (0)			(1) (0) (0)			
S. VOLUME (DECREASE)			(2) (0) (0)			(1) (0) (0)			
S. VOLUME (INCREASE)			(5) (0) (0)			(1) (0) (0)			
ABS. LIMIT			(0) () (INCREASE)						
PROZONE LIMIT			(0) () (LOWER)						
REAGENT									
R1			(300) (0) (*) (0)						
R2			(0) (0) (*) (0)						
R3			(0) (0) (*) (0)						
R4			(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT			(0.1)			
SPAN (0)			DUPLICATE LIMITE			(400)			
2 POINT (0)			SENSITIVITY LIMIT			(0)			
FULL (0)			S1 ABS LIMIT			(-32000) (32000)			
CHANGE LOT (CANCEL)			COMPENSATED LIMIT			()			
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter expected values.



Modification from the previous version

(06/2009)
AA-H911-CHSL-5



CHOLESTEROL

☞ Ref. : CHOL-0220 12 x 20 mL
 CHOL-0420 9 x 50 mL
 CHOL-0520 6 x 100 mL

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (CHOL) (*)		TEST NAME (CHOL)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (CHOLESTEROL)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(5)		(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CHOL)									
ASSAY CODE (1 POINT)(10)()			WAVELENGTH (SUB/MAIN) (700)/(505)						
ASSAY POINT (31) (0) (0) (0)			DILUTION (00301) (99)						
			< S.TYPE 1 >		< S.TYPE 2 >				
S. VOLUME (REGULAR)			(3) (0) (0)		(1) (0) (0)				
S. VOLUME (DECREASE)			(2) (0) (0)		(1) (0) (0)				
S. VOLUME (INCREASE)			(5) (0) (0)		(1) (0) (0)				
ABS. LIMIT			(0) () (INCREASE)						
PROZONE LIMIT			(0) () (LOWER)						
REAGENT			(300) (0) (*) (0)						
R1			(0) (0) (*) (0)						
R2			(0) (0) (*) (0)						
R3			(0) (0) (*) (0)						
R4			(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT (0.1)						
SPAN (0)			DUPLICATE LIMITE (400)						
2 POINT (0)			SENSITIVITY LIMIT (0)						
FULL (0)			S1 ABS LIMIT (-32000) (32000)						
CHANGE LOT (CANCEL)			COMPENSATED LIMIT ()						
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.

☞ Modification from the previous version

(06/2009)
 AA-H911-CHOL-4



CHOLESTEROL HDL

Ref. : HDLC-0060 3 x 10 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use : 


Working temperature : 37°C

TEST	(HDL) (*)	TEST NAME	(HDL)	UNITS	(g/l)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(HDL CHOLESTEROL)						
CONTROL INTERVAL	(300)	INSTRUM. FACTOR	(Y=aX+b)	a (1.0) b (0)					
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	AGE M F								
	(0) (Y) (0) (0) (0) (0)		(0) (0)						
	(0) (Y) (0) (0) (§§) (§§)		(0) (0) (§§) (§§)						
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(5)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(15)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(15)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (HDL)		WAVELENGTH (SUB/MAIN)	(700)/(505)						
ASSAY CODE	(1 POINT)(10)()	DILUTION	(00301) (99)						
ASSAY POINT	(31) (0) (0) (0)								
S. VOLUME (REGULAR)		< S.TYPE 1 >	< S.TYPE 2 >						
S. VOLUME (DECREASE)		(15) (0) (0)	(1) (0) (0)						
S. VOLUME (INCREASE)		(7) (0) (0)	(1) (0) (0)						
ABS. LIMIT		(30) (0) (0)	(1) (0) (0)						
PROZONE LIMIT		(0) () (INCREASE)							
REAGENT	R1	(0) () (LOWER)							
	R2	(300) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(0.1)						
SPAN	(0)	DUPLICATE LIMITE	(400)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(-32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter expected values.

 Modification from the previous version



CHOLESTEROL HDL SL 2G

Ref. HDLL-0380 1 x 80 mL
HDLL-0390 3 x 80 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use : 

Working temperature : 37°C

TEST	(HDL) (*)	TEST NAME	(HDL)	UNITS	(mg/dL)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(HDL CHOL.)						
CONTROL INTERVAL	(300)	INSTRUM. FACTOR (Y=aX+b)	a (1.0) b (0)						
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
AGE	M F								
(0) (Y)	(0) (0) (0) (0)		(0) (0)						
(0) (Y)	(0) (0) (0) (0)		(0) (0)						
	(§§) (§§) (§§) (§§)								
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(6)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (HDL)		WAVELENGTH (SUB/MAIN)	(700)/(600)						
ASSAY CODE (2:2POINT-END)	(10)()	DILUTION (*)	(0)						
ASSAY POINT (15) (31) () ()		< S.TYPE 1 >	< S.TYPE 2 >						
S. VOLUME (REGULAR)		(3) (0) (0)	(0) (0) (0)						
S. VOLUME (DECREASE)		(3) (0) (0)	(0) (0) (0)						
S. VOLUME (INCREASE)		(3) (0) (0)	(0) (0) (0)						
ABS. LIMIT		(0) 32000) (2 : INCREASE)							
PROZONE LIMIT		(0) (0) (2 : LOWER)							
REAGENT R1		(300) (0) (*) (0)							
R2		(0) (0) (*) (0)							
R3		(100) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(0.1)						
SPAN	(0)	DUPLICATE LIMITE	(1000)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(-32000) (32000)						
BOTTLE	(CANCEL)								

*: Enter position or code number
§§: Enter expected values
#: User defined

*** : Enter calibrator value



CHOLESTEROL LDL SL 2G

Ref. LDLL-0380 1 x 80 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use : 

Working temperature : 37°C

TEST	(LDL) (*)	TEST NAME	(LDL)	UNITS	(mg/dL)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(LDL CHOL.)						
CONTROL INTERVAL	(300)	INSTRUM. FACTOR	(Y=aX+b)	a (1.0)	b (0)				
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
AGE	M F								
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(6)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (LDL)	ASSAY CODE (2:2POINT-END)(10)()		WAVELENGTH (SUB/MAIN) (660)/(546)						
	ASSAY POINT (16) (31) (0) (0)		DILUTION (*) (0)						
			< S.TYPE 1 >			< S.TYPE 2 >			
S. VOLUME (REGULAR)			(3) (0) (0)			(0) (0) (0)			
S. VOLUME (DECREASE)			(3) (0) (0)			(0) (0) (0)			
S. VOLUME (INCREASE)			(3) (0) (0)			(0) (0) (0)			
ABS. LIMIT			(0) (0) (0) (2 : INCREASE)						
PROZONE LIMIT			(0) (0) (2 : LOWER)						
REAGENT	R1			(300) (0) (*) (0)					
	R2			(0) (0) (*) (0)					
	R3			(100) (0) (*) (0)					
	R4			(0) (0) (*) (0)					
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK	(0)			STD LIMIT			(0.1)		
SPAN	(0)			DUPLICATE LIMITE			(1000)		
2 POINT	(0)			SENSITIVITY LIMIT			(0)		
FULL	(0)			S1 ABS LIMIT			(-32000) (32000)		
BOTTLE	(CANCEL)								

*: Enter position or code number
§§: Enter expected values
#: User defined

*** : Enter calibrator value



CHOLINESTERASE

Ref.: CHES-0053 16 x 3 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (CHE) (*)		TEST NAME (CHE)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (CHOLINESTERASE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(5000)		(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (ALP)									
ASSAY CODE (KIN A)(10)()			WAVELENGTH (SUB/MAIN) (700)/(415)						
ASSAY POINT (19) (23) (0) (0)			DILUTION (00301) (99)						
			< S.TYPE 1 >		< S.TYPE 2 >				
S. VOLUME (REGULAR)			(3) (0) (0)		(1) (0) (0)				
S. VOLUME (DECREASE)			(1) (0) (0)		(1) (0) (0)				
S. VOLUME (INCREASE)			(6) (0) (0)		(1) (0) (0)				
ABS. LIMIT			(8000) () (2 : INCREASE)						
PROZONE LIMIT			(0) () (2 : LOWER)						
REAGENT		R1	(300) (0) (*) (0)						
		R2	(0) (0) (*) (0)						
		R3	(25) (0) (*) (0)						
		R4	(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK		(0)	STD LIMIT		(0.1)				
SPAN		(0)	DUPLICATE LIMITE		(200)				
2 POINT		(0)	SENSITIVITY LIMIT		(0)				
FULL		(0)	S1 ABS LIMIT		(-32000) (32000)				
CHANGE LOT		(CANCEL)	COMPENSATED LIMIT		()				
BOTTLE		(CANCEL)							

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter values.

Modification from the previous version

(06/2009)
AA-H911-CHES-5



CK NAC SL

ONE REAGENT PROCEDURE

Ref. : CKSL-0410 2 x 62.5 mL
 CKSL-0430 4 x 62.5 mL

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (CPK) (*)		TEST NAME (CPK)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (CPK)							
CONTROL INTERVAL (500)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
AGE M F									
(0) (Y) (0) (0) (0) (0)		(0) (0)							
(0) (Y) (0) (0) (0) (0)		(0) (0)							
		(\$\$) (\$\$)		(\$\$) (\$\$)					
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(1200)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CPK)		ASSAY CODE (6 KINETIC A)(10)()		WAVELENGTH (SUB/MAIN) (415)/(340)					
ASSAY POINT (9) (20) (0) (0)				DILUTION (00301) (99)					
S. VOLUME (REGULAR)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (DECREASE)		(7) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(3) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(14) (0) (0)		(1) (0) (0)					
PROZONE LIMIT		(9000) () (2 : INCREASE)							
REAGENT R1		(0) () (2 : LOWER)							
R2		(300) (0) (*) (*)							
R3		(0) (0) (*) (*)							
R4		(0) (0) (*) (*)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(100)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (NON)		COMPENSATED LIMIT		()					
BOTTLE (NON)									

* : Enter position or code.
 \$\$\$: Enter the expected values.

*** : Enter calibrator or standard value.



CK NAC

Ref. : CKNA-0030 20 x 3 mL

**APPLICATION HITACHI 911
PROPOSAL**


Instruction use : 

Working temperature : 37°C

TEST (CPK) (*)		TEST NAME (CPK)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (CPK)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
AGE M F									
(0) (Y) (0) (0) (0) (0)		(0) (0)							
(0) (Y) (0) (0) (0) (0)		(0) (0)							
		(§§) (§§)		(§§) (§§)					
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(1500)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(**)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CPK)									
ASSAY CODE (KIN A)(10)()						WAVELENGTH (SUB/MAIN) (415)/(340)			
ASSAY POINT (9) (20) (0) (0)						DILUTION (00301) (99)			
						< S.TYPE 1 >		< S.TYPE 2 >	
S. VOLUME (REGULAR)						(7) (0) (0)		(1) (0) (0)	
S. VOLUME (DECREASE)						(3) (0) (0)		(1) (0) (0)	
S. VOLUME (INCREASE)						(14) (0) (0)		(1) (0) (0)	
ABS. LIMIT						(9000) () (2 : INCREASE)			
PROZONE LIMIT						(0) () (2 : LOWER)			
REAGENT R1						(300) (0) (*) (0)			
R2						(0) (0) (*) (0)			
R3						(0) (0) (*) (0)			
R4						(0) (0) (*) (0)			
CALIBRATION TYPE						(LINEAR) (2) (2) (0) ()			
AUTO CALIB.									
TIME OUT BLANK (0)						STD LIMIT (0.1)			
SPAN (0)						DUPLICATE LIMITE (100)			
2 POINT (0)						SENSITIVITY LIMIT (0)			
FULL (0)						S1 ABS LIMIT (-32000) (32000)			
CHANGE LOT (CANCEL)						COMPENSATED LIMIT ()			
BOTTLE (CANCEL)									

* : Enter position or code number.
 §§§ : Enter values.

*** : Enter calibrator or standard value.

 Modification from the previous version

(06/2009)
AA-H911-CKNA-4



CK - MB SL

Ref.: CMSL-0410 2 x 62.5 mL
 CMSL-0430 4 x 62.5 mL

ONE REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (CKMB) (*)		TEST NAME (CKMB)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (CKMB)							
CONTROL INTERVAL (500)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(1200)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CKMB)		WAVELENGTH (SUB/MAIN) (415)/(340)							
ASSAY CODE (6 KINETIC A)(10)()		DILUTION (00301) (99)							
ASSAY POINT (9) (20) (0) (0)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(7) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(14) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(9000) () (2 : INCREASE)							
PROZONE LIMIT		(0) () (2 : LOWER)							
REAGENT R1		(300) (0) (*) (*)							
	R2	(0) (0) (*) (*)							
	R3	(0) (0) (*) (*)							
	R4	(0) (0) (*) (*)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(100)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (NON)		COMPENSATED LIMIT		()					
BOTTLE (NON)									

* : Enter position or code.
 §§§ : Enter the expected values.

*** : Enter calibrator or standard value.



CK - MB

Ref. : CKMB-0030 20 x 3 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (CKMB) (*)		TEST NAME (CKMB)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (CKMB)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(1500)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(**)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CKMB)		ASSAY CODE (KIN A)(10)()		WAVELENGTH (SUB/MAIN) (415)/(340)					
		ASSAY POINT (9) (20) (0) (0)		DILUTION (00301) (99)					
				< S.TYPE 1 >					
S. VOLUME (REGULAR)				(15) (0) (0)					
S. VOLUME (DECREASE)				(7) (0) (0)					
S. VOLUME (INCREASE)				(20) (0) (0)					
ABS. LIMIT				(9000) () (2 : INCREASE)					
PROZONE LIMIT				(0) () (2 : LOWER)					
REAGENT		R1		(300) (0) (*) (0)					
		R2		(0) (0) (*) (0)					
		R3		(0) (0) (*) (0)					
		R4		(0) (0) (*) (0)					
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(100)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter values.



COPPER

Ref. : CUIV-0050 5 x 10 mL

ONE REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (CU) (*)		TEST NAME (CU)		UNITS (µg/dl)					
DATA MODE (1: ON BOARD)		REPORT NAME (COPPER)							
CONTROL INTERVAL (80)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)	(0) (0)					
	(0) (Y)	(0) (0)	(0) (0)						
		(§§) (§§)	(§§) (§§)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(500)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(15)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(15)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CU)		WAVELENGTH (SUB/MAIN) (700)/(570)							
ASSAY CODE (1 POINT)(10)()		DILUTION (00301) (99)							
ASSAY POINT (31) (0) (0) (0)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(15) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(10) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(20) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(0) () (LOWER)							
REAGENT		(300) (0) (*) (0)							
	R1	(0) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (12)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(400)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter expected values.



COPPER

Ref. : CUIV-0050 5 x 10 mL

TWO REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (CU) (*)		TEST NAME (CU)		UNITS (µg/dl)					
DATA MODE (1: ON BOARD)		REPORT NAME (COPPER)							
CONTROL INTERVAL (80)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)	(0) (0)					
	(0) (Y)	(0) (0)	(0) (0)						
		(§§) (§§)	(§§) (§§)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(500)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(15)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(15)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(15)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CU)		WAVELENGTH (SUB/MAIN) (700)/(570)							
ASSAY CODE (1 POINT)(10)()		DILUTION (00301) (99)							
ASSAY POINT (31) (0) (0) (0)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(15) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(7) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(30) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(0) () (LOWER)							
REAGENT		(300) (0) (*) (0)							
	R1	(15) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (12)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(400)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.




CREATININE JAFFE

Ref. : CRCO-0600 2 x 125 mL
CRCO-0700 4 x 250 mL

ONE REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use : 

Working temperature : 37°C

TEST	(CREA) (*)	TEST NAME	(CREA)	UNITS	(mg/L)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(CREATININE JAFFE)						
CONTROL INTERVAL	(80)	INSTRUM. FACTOR	(Y=aX+b)	a (1.0) b (0)					
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	AGE M F								
	(0) (Y) (0) (0) (0) (0)				(0) (0)				
	(0) (Y) (0) (0) (0) (0)								
	(§§) (§§) (§§) (§§)								
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(50)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CREA)		WAVELENGTH (SUB/MAIN)	(600)/(505)						
ASSAY CODE	(1 POINT)(10)()	DILUTION	(00301) (99)						
ASSAY POINT	(31) (0) (0) (0)								
		< S.TYPE 1 >	< S.TYPE 2 >						
S. VOLUME (REGULAR)		(20) (0) (0)	(1) (0) (0)						
S. VOLUME (DECREASE)		(5) (0) (0)	(1) (0) (0)						
S. VOLUME (INCREASE)		(40) (0) (0)	(1) (0) (0)						
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(32000) () (UPPER)							
REAGENT	R1	(250) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(0.1)						
SPAN	(0)	DUPLICATE LIMITE	(200)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(-32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter expected values.



CREATININE PAP SL

Ref.: CRSL-0630 2 x 133 mL
 CRSL-0637 2 x 133 mL + STD

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (CRSL) (*)		TEST NAME (CRSL)		UNITS (mg/dL)					
DATA MODE (1: ON BOARD)		REPORT NAME (CREAT)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)			(0) (0)				
(0) (Y)	(0) (0)	(0) (0)							
	(\$\$) (\$\$)	(\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0.3)-(30)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(8)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(8)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (CRSL)		WAVELENGTH (SUB/MAIN) (...)/(546)							
ASSAY CODE (KIN A)(10)()		DILUTION (00301) (99)							
ASSAY POINT (19) (25) (0) (0)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(8) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(4) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(16) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(xxx) () (2 : INCREASE)							
PROZONE LIMIT		(0) () (1 : HIGHER)							
REAGENT									
	R1	(210) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(70) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number
 \$\$\$: Enter expected values.
 xxx : Enter data by the user

***: Enter calibrator or standard value

GAMMA GT SL

ONE REAGENT PROCEDURE

Ref. : GASL-0400 2 x 62.5 mL
 GASL-0420 4 x 62.5 mL
 GASL-0500 5 x 125 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (GGT) (*)		TEST NAME (GGT)		UNITS (IU/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (GGT SL ONE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(300)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (GGT)			WAVELENGTH (SUB/MAIN) (700)/(415)						
ASSAY CODE (KIN A)(10)()			DILUTION (00301) (99)						
ASSAY POINT (6) (15) (0) (0)									
			< S.TYPE 1 >	< S.TYPE 2 >					
S. VOLUME (REGULAR)			(20) (0) (0)	(1) (0) (0)					
S. VOLUME (DECREASE)			(3) (0) (0)	(1) (0) (0)					
S. VOLUME (INCREASE)			(40) (0) (0)	(1) (0) (0)					
ABS. LIMIT			(8000) () (2 : INCREASE)						
PROZONE LIMIT			(0) () (2 : LOWER)						
REAGENT			(250) (0) (*) (0)						
R1			(0) (0) (*) (0)						
R2			(0) (0) (*) (0)						
R3			(0) (0) (*) (0)						
R4			(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT		(0.1)				
SPAN (0)			DUPLICATE LIMITE		(200)				
2 POINT (0)			SENSITIVITY LIMIT		(0)				
FULL (0)			S1 ABS LIMIT		(-32000) (32000)				
CHANGE LOT (CANCEL)			COMPENSATED LIMIT		()				
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter values.



GAMMA GT SL

Ref. : GASL-0400 2 x 62.5 mL
 GASL-0420 4 x 62.5 mL
 GASL-0500 5 x 125 mL

TWO REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (GGT) (*)		TEST NAME (GGT)		UNITS (IU/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (GGT SL TWO)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(300)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (GGT)		WAVELENGTH (SUB/MAIN) (700)/(415)							
ASSAY CODE (KIN A)(10)()		DILUTION (00301) (99)							
ASSAY POINT (19) (31) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(20) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(40) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(8000) () (2 : INCREASE)							
PROZONE LIMIT		(0) () (2 : LOWER)							
REAGENT R1		(200) (0) (*) (0)							
		R2							
		(0) (0) (*) (0)							
		R3							
		(50) (0) (*) (0)							
		R4							
		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter values.



GAMMA GT

Ref. : GAGT-0030 20 x 3 mL
 GAGT-0200 12 x 20 mL

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (GGT) (*)		TEST NAME (GGT)		UNITS (IU/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (GGT)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(300)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(20)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(20)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (GGT)		WAVELENGTH (SUB/MAIN) (700)/(415)							
ASSAY CODE (KIN A)(10)()		DILUTION (00301) (99)							
ASSAY POINT (6) (15) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(20) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(40) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(8000) () (2 : INCREASE)							
PROZONE LIMIT		(0) () (2 : LOWER)							
REAGENT R1		(250) (0) (*) (0)							
R2		(0) (0) (*) (0)							
R3		(0) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter values.



GLUCOSE HK SL

Ref. : GHSL-0600 5 x 125 mL

ONE REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (GLU) (*)		TEST NAME (GLU)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (GLUCOSE HK SL ONE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
AGE M F									
(0) (Y) (0) (0) (0) (0)		(0) (0)							
(0) (Y) (0) (0) (§§) (§§)		(0) (0) (§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(6)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(**)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (GLU)		WAVELENGTH (SUB/MAIN) (380)/(340)							
ASSAY CODE (1 POINT)(10)()		DILUTION (00301) (99)							
ASSAY POINT (31) (0) (0) (0)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(2) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(6) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(0) () (LOWER)							
REAGENT		(300) (0) (*) (0)							
R1		(0) (0) (*) (0)							
R2		(0) (0) (*) (0)							
R3		(0) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT (0.1)							
SPAN (0)		DUPLICATE LIMITE (200)							
2 POINT (0)		SENSITIVITY LIMIT (0)							
FULL (0)		S1 ABS LIMIT (-32000) (32000)							
CHANGE LOT (CANCEL)		COMPENSATED LIMIT ()							
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter expected values.



GLUCOSE HK SL

Ref. : GHSL-0600 5 x 125 mL

TWO REAGENT PROCEDURE

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (GLU) (*)		TEST NAME (GLU)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (GLUCOSE HK SL TWO)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(6)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (GLU)		WAVELENGTH (SUB/MAIN) (380)/(340)							
ASSAY CODE (2 POINT END)(5)()		DILUTION (00301) (99)							
ASSAY POINT (4) (15) (0) (0)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(2) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(6) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(0) () (LOWER)							
REAGENT		(200) (0) (*) (0)							
	R1	(50) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



GLUCOSE PAP SL

Ref. : GPSL-0490	1 x 100 mL
GPSL-0500	6 x 100 mL
GPSL-0700	4 x 250 mL
GPSL-0507	6 x 100 mL + STD
GPSL-0707	4 x 250 mL + STD

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (GLU) (*)	TEST NAME (GLU)	UNITS (g/l)
DATA MODE (1: ON BOARD)	REPORT NAME (GLUCOSE SL)	
CONTROL INTERVAL (300)	INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)	
EXPECTED VALUE <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>	
AGE M F		
(0) (Y) (0) (0) (0) (0)		(0) (0)
(0) (Y) (0) (0) (0) (0)		
(§§) (§§) (§§) (§§)		
TECHNICAL LIMIT <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>	
(0)-(4)	(0)-(0)	
STD (1)	CONC (0)	POS (*)
(2)	(***)	(*)
(3-6)	(0)	(0)
S. VOL (3)	PRE. DIL. (0)	VOL (0)
(0)	(0)	(*)
(0)	(0)	(*)
(0)	(0)	(*)
LOT (1)	QUALITATIVE (0)	(NO) ()
(2)	(0)	()
(3-6)	(0)	()

TEST (GLU)	WAVELENGTH (SUB/MAIN) (700)/(505)
ASSAY CODE (2 POINT/CIN)(10)()	DILUTION (00301) (99)
ASSAY POINT (6) (18) (0) (0)	
S. VOLUME (REGULAR)	< S.TYPE 1 > (3) (0) (0)
S. VOLUME (DECREASE)	< S.TYPE 2 > (1) (0) (0)
S. VOLUME (INCREASE)	(2) (0) (0)
ABS. LIMIT	(6) (0) (0)
PROZONE LIMIT	(15000) () (2 : INCREASE)
REAGENT R1	(32000) () (2 : HIGHER)
R2	(300) (0) (*) (0)
R3	(0) (0) (*) (0)
R4	(0) (0) (*) (0)
CALIBRATION TYPE	(0) (0) (*) (0)
AUTO CALIB.	(LINEAR) (2) (2) (0) ()
TIME OUT BLANK (0)	STD LIMIT (0.1)
SPAN (0)	DUPLICATE LIMITE (1100)
2 POINT (0)	SENSITIVITY LIMIT (0)
FULL (0)	S1 ABS LIMIT (-32000) (32000)
CHANGE LOT (CANCEL)	COMPENSATED LIMIT ()
BOTTLE (CANCEL)	

* : Enter position or code number.
 §§§ : Enter values.

*** : Enter calibrator or standard value.

Modification from the previous version

(06/2009)
AA-H911-GPSL-5



GLUCOSE PAP

Ref. : GLUP-0700 4 x 250 mL
 GLUP-0800 5 x 500 mL

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST	(GLU) (*)	TEST NAME	(GLU)	UNITS	(g/l)								
DATA MODE	(1: ON BOARD)	REPORT NAME	(GLUCOSE)										
CONTROL INTERVAL	(300)	INSTRUM. FACTOR	(Y=aX+b)	a	(1.0) b (0)								
EXPECTED VALUE	<S.TYPE 1>		EXPECTED VALUE	<S.TYPE 2>									
	AGE	M	F										
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)								
	(0) (Y)	(0) (0)	(0) (0)										
		(\$\$) (\$\$)	(\$\$) (\$\$)										
TECHNICAL LIMIT	<S.TYPE 1>		EXPECTED VALUE	<S.TYPE 2>									
	(0)-(4)			(0)-(0)									
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)				
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()				
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()				
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()				
TEST (GLU)													
ASSAY CODE			(2 POINT/CIN)(10) ()			WAVELENGTH (SUB/MAIN)				(700)/(505)			
ASSAY POINT			(6) (18) (0) (0)			DILUTION				(00301) (99)			
S. VOLUME (REGULAR)						< S.TYPE 1 >		< S.TYPE 2 >					
						(3) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)						(2) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)						(6) (0) (0)		(1) (0) (0)					
ABS. LIMIT						(15000) () (2 : INCREASE)							
PROZONE LIMIT						(32000) () (1 : HIGHER)							
REAGENT						R1		(300) (0) (*) (0)					
						R2		(0) (0) (*) (0)					
						R3		(0) (0) (*) (0)					
						R4		(0) (0) (*) (0)					
CALIBRATION TYPE						(LINEAR) (2) (2) (0) ()							
AUTO CALIB.													
TIME OUT BLANK						(0)		STD LIMIT				(0.1)	
SPAN						(0)		DUPLICATE LIMITE				(1100)	
2 POINT						(0)		SENSITIVITY LIMIT				(0)	
FULL						(0)		S1 ABS LIMIT				(-32000) (32000)	
CHANGE LOT						(CANCEL)		COMPENSATED LIMIT				()	
BOTTLE						(CANCEL)							

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter values.



IRON FERROZINE®

Ref. : FEFR-0600

2 x 125 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (IRON) (*)		TEST NAME (IRON)		UNITS (mg/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (IRON FERROZINE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(5)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(25)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(25)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (IRON)		ASSAY CODE (2 POINTEND)(10)()		WAVELENGTH (SUB/MAIN) (700)/(570)					
ASSAY POINT (15) (31) (0) (0)				DILUTION (00301) (99)					
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(25) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(10) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(40) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (2 : INCREASE)							
PROZONE LIMIT		(32000) () (1 : HIGHER)							
REAGENT		(250) (0) (*) (0)							
R1		(0) (0) (*) (0)							
R2		(25) (0) (*) (0)							
R3		(0) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(1100)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter values.



LDH-P 4 + 1 SL

Réf. : LDSL-0410 2 x 62.5 mL
 LDSL-0430 4 x 62.5 mL

ONE REAGENT PROCEDURE

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (LDH) (*)		TEST NAME (LDH)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (LDH SL ONE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y= aX+ b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)				(0) (0)					
(0) (Y) (0) (0) (§§) (§§)		(0) (0) (§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(1500)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (LDH)									
ASSAY CODE (KIN A)(10)()						WAVELENGTH (SUB/MAIN) (415)/(340)			
ASSAY POINT (7) (17) (0) (0)						DILUTION (00301) (99)			
						< S.TYPE 1 >		< S.TYPE 2 >	
S. VOLUME (REGULAR)						(7) (0) (0)		(1) (0) (0)	
S. VOLUME (DECREASE)						(4) (0) (0)		(1) (0) (0)	
S. VOLUME (INCREASE)						(17) (0) (0)		(1) (0) (0)	
ABS. LIMIT						(4000) () (1 : DECREASE)			
PROZONE LIMIT						(0) () (2 : LOWER)			
REAGENT R1						(250) (0) (*) (0)			
R2						(0) (0) (*) (0)			
R3						(0) (0) (*) (0)			
R4						(0) (0) (*) (0)			
CALIBRATION TYPE						(LINEAR) (2) (2) (0) ()			
AUTO CALIB.									
TIME OUT BLANK (0)						STD LIMIT (0.1)			
SPAN (0)						DUPLICATE LIMITE (100)			
2 POINT (0)						SENSITIVITY LIMIT (0)			
FULL (0)						S1 ABS LIMIT (-32000) (32000)			
CHANGE LOT (CANCEL)						COMPENSATED LIMIT ()			
BOTTLE (CANCEL)									

* : Enter position or code number
 §§§ : Enter values

*** : Enter calibrator or standard value



LDH-P 4 + 1 SL

Réf. : LDSL-0410 2 x 62.5 mL
 LDSL-0430 4 x 62.5 mL

TWO REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (LDH) (*)		TEST NAME (LDH)		UNITS (U/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (LDH SL TWO)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)				(0) (0)					
(0) (Y) (0) (0) (0) (0)									
		(§§) (§§) (§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(1500)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(6)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(6)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (LDH)									
ASSAY CODE (KIN A)(10)()						WAVELENGTH (SUB/MAIN) (415)/(340)			
ASSAY POINT (19) (31) (0) (0)						DILUTION (00301) (99)			
						< S.TYPE 1 >		< S.TYPE 2 >	
S. VOLUME (REGULAR)						(4) (0) (0)		(1) (0) (0)	
S. VOLUME (DECREASE)						(3) (0) (0)		(1) (0) (0)	
S. VOLUME (INCREASE)						(17) (0) (0)		(1) (0) (0)	
ABS. LIMIT						(4000) () (1 : DECREASE)			
PROZONE LIMIT						(0) () (2 : LOWER)			
REAGENT R1						(200) (0) (*) (0)			
R2						(0) (0) (*) (0)			
R3						(50) (0) (*) (0)			
R4						(0) (0) (*) (0)			
CALIBRATION TYPE						(LINEAR) (2) (2) (0) ()			
AUTO CALIB.									
TIME OUT BLANK (0)						STD LIMIT (0.1)			
SPAN (0)						DUPLICATE LIMITE (100)			
2 POINT (0)						SENSITIVITY LIMIT (0)			
FULL (0)						S1 ABS LIMIT (-32000) (32000)			
CHANGE LOT (CANCEL)						COMPENSATED LIMIT ()			
BOTTLE (CANCEL)									

* : Enter position or code number
 §§§ : Enter values

*** : Enter calibrator or standard value



LDH - P

Ref.: LDHP-0030 20 x 3 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use : 

Working temperature : 37°C


TEST (LDH) (*)	TEST NAME (LDH)	UNITS (U/L)
DATA MODE (1: ON BOARD)	REPORT NAME (LDH)	
CONTROL INTERVAL (300)	INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)	
EXPECTED VALUE <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>	
AGE M F		
(0) (Y) (0) (0) (0) (0)		(0) (0)
(0) (Y) (0) (0) (0) (0)		
(§§) (§§) (§§) (§§)		
TECHNICAL LIMIT <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>	
	(0)-(1500)	(0)-(0)
STD (1)	CONC (0)	POS (*)
(2)	(**)	(*)
(3-6)	(0)	(0)
S. VOL (7)	PRE. DIL. (0)	VOL (0)
		CODE (*)
		LOT (1)
		QUALITATIVE (0)
		(NO) ()

TEST (LDH)	WAVELENGTH (SUB/MAIN) (415)/(340)
ASSAY CODE (KIN A)(10)()	DILUTION (00301) (99)
ASSAY POINT (7) (17) (0) (0)	
	< S.TYPE 1 >
S. VOLUME (REGULAR)	(7) (0) (0)
S. VOLUME (DECREASE)	(3) (0) (0)
S. VOLUME (INCREASE)	(17) (0) (0)
ABS. LIMIT	(4000) () (1 : DECREASE)
PROZONE LIMIT	(0) () (2 : LOWER)
REAGENT R1	(300) (0) (*) (0)
R2	(0) (0) (*) (0)
R3	(0) (0) (*) (0)
R4	(0) (0) (*) (0)
CALIBRATION TYPE	(LINEAR) (2) (2) (0) ()
AUTO CALIB.	
TIME OUT BLANK (0)	STD LIMIT (0.1)
SPAN (0)	DUPLICATE LIMITE (100)
2 POINT (0)	SENSITIVITY LIMIT (0)
FULL (0)	S1 ABS LIMIT (-32000) (32000)
CHANGE LOT (CANCEL)	COMPENSATED LIMIT ()
BOTTLE (CANCEL)	

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter values.

 Modification from the previous version


(06/2009)
AA-H911-LDHP-4



MAGNESIUM CALMAGITE

Ref. : MAGN-0600 2 x 125 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use : 

Working temperature : 37°C

TEST	(Mg) (*)	TEST NAME	(Mg)	UNITS	(mg/L)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(MAGNESIUM)						
CONTROL INTERVAL	(300)	INSTRUM. FACTOR	(Y=aX+b)	a (1.0)	b (0)				
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	AGE M F								
	(0) (Y) (0) (0) (0) (0)		(0) (0)						
	(0) (Y) (0) (0) (0) (0)								
	(\$\$) (\$\$)		(\$\$) (\$\$)						
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE	<S.TYPE 2>						
	(0)-(-50)		(0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(4)	(0)	(0)	(*)	(1)	(0)	()
(2)	(**)	(*)	(4)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST	(Mg)	WAVELENGTH	(SUB/MAIN)	(600)/(505)					
ASSAY CODE	(1 POINT)(10)()	DILUTION	(00301) (99)						
ASSAY POINT	(31) (0) (0) (0)								
S. VOLUME (REGULAR)		< S.TYPE 1 >	< S.TYPE 2 >						
S. VOLUME (DECREASE)		(4) (0) (0)	(1) (0) (0)						
S. VOLUME (INCREASE)		(2) (0) (0)	(1) (0) (0)						
ABS. LIMIT		(8) (0) (0)	(1) (0) (0)						
PROZONE LIMIT		(0) () (2 : INCREASE)							
REAGENT	R1	(0) () (2 : LOWER)							
	R2	(200) (0) (*) (0)							
	R3	(200) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(0.1)						
SPAN	(0)	DUPLICATE LIMITE	(2500)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(-32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter values.



MICROPROTEIN

Ref. : PRTP-0600

2 x 125 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (MPRO) (*)		TEST NAME (MPRO)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (MICROPROTEIN)							
CONTROL INTERVAL (80)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(\$\$) (\$\$)	(\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(6)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(6)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (MPRO)		ASSAY CODE (1 POINT)(10)()		WAVELENGTH (SUB/MAIN) (700)/(600)					
ASSAY POINT (31) (0) (0) (0)		DILUTION (00301) (99)							
S. VOLUME (REGULAR)			< S.TYPE 1 >		< S.TYPE 2 >				
			(6) (0) (0)		(1) (0) (0)				
S. VOLUME (DECREASE)			(3) (0) (0)		(1) (0) (0)				
S. VOLUME (INCREASE)			(10) (0) (0)		(1) (0) (0)				
ABS. LIMIT			(0) () (INCREASE)						
PROZONE LIMIT			(32000) () (UPPER)						
REAGENT		R1	(300) (0) (*) (0)						
		R2	(0) (0) (*) (0)						
		R3	(0) (0) (*) (0)						
		R4	(0) (0) (*) (0)						
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK		(0)	STD LIMIT		(0.1)				
SPAN		(0)	DUPLICATE LIMITE		(300)				
2 POINT		(0)	SENSITIVITY LIMIT		(0)				
FULL		(0)	S1 ABS LIMIT		(-32000) (32000)				
CHANGE LOT		(CANCEL)	COMPENSATED LIMIT		()				
BOTTLE		(CANCEL)							

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



PHOSPHORUS

Ref. : PHOS-0600 2 x 125 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (PHOS) (*)		TEST NAME (PHOS)		UNITS (mg/L)					
DATA MODE (1: ON BOARD)		REPORT NAME (PHOSPHORE)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(200)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (PHOS)									
ASSAY CODE (1 POINT)(10)()			WAVELENGTH (SUB/MAIN) (700)/(340)						
ASSAY POINT (31) (0) (0) (0)			DILUTION (00301) (99)						
			< S.TYPE 1 >			< S.TYPE 2 >			
S. VOLUME (REGULAR)			(3) (0) (0)			(1) (0) (0)			
S. VOLUME (DECREASE)			(2) (0) (0)			(1) (0) (0)			
S. VOLUME (INCREASE)			(6) (0) (0)			(1) (0) (0)			
ABS. LIMIT			(0) () (2 : INCREASE)						
PROZONE LIMIT			(32000) () (1 : UPPER)						
REAGENT R1			(300) (0) (*) (0)						
			(0) (0) (*) (0)						
			(0) (0) (*) (0)						
			(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT			(0.1)			
SPAN (0)			DUPLICATE LIMITE			(600)			
2 POINT (0)			SENSITIVITY LIMIT			(0)			
FULL (0)			S1 ABS LIMIT			(-32000) (32000)			
CHANGE LOT (CANCEL)			COMPENSATED LIMIT			()			
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter values.



TOTAL PROTEIN

Ref. : PRTB-0600 2 x 125 mL
 PRTB-0700 4 x 250 mL

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (PROT) (*)		TEST NAME (PROT)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (PROTEIN BIURET)							
CONTROL INTERVAL (80)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(\$\$) (\$\$)	(\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(100)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (PROT)		ASSAY CODE (1 POINT)(10)()		WAVELENGTH (SUB/MAIN) (700)/(546)					
ASSAY POINT (31) (0) (0) (0)				DILUTION (00301) (99)					
S. VOLUME (REGULAR)		< S.TYPE 1 >		< S.TYPE 2 >					
		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(1) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(6) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(32000) () (UPPER)							
REAGENT		R1		(300) (0) (*) (0)					
		R2		(0) (0) (*) (0)					
		R3		(0) (0) (*) (0)					
		R4		(0) (0) (*) (0)					
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(300)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



TRIGLYCERIDES MONO SL NEW

Ref. : TGML-0425	6 x 50 mL
TGML-0515	6 x 100 mL
TGML-0700	4 x 250 mL
☞ TGML-0427	6 x 50 mL + STD
☞ TGML-0517	6 x 100 mL + STD
☞ TGML-0707	4 x 250 mL + STD

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :
 Working temperature : 37°C

TEST (TG) (*)	TEST NAME (TG)	UNITS (g/l)	
DATA MODE (1: ON BOARD)	REPORT NAME (TRIGLYCERIDES)		
CONTROL INTERVAL (300)	INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)		
EXPECTED VALUE <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>		
AGE M F			
(0) (Y) (0) (0) (0) (0)			(0) (0)
(0) (Y) (0) (0) (0) (0)			
(\$\$) (\$\$) (\$\$) (\$\$)			
TECHNICAL LIMIT <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>		
(0)-(10)	(0)-(0)		
STD (1) (0)	CONC (2) (**)	POS (*)	S. VOL (3) (0)
(3-6) (0)	(0)	(0)	(0)
			PRE. DIL. (0) (0)
			VOL (0) (0)
			CODE (*)
			LOT (1) (2) (3-6)
			QUALITATIVE (0) (0) (0)
			(NO) () () ()
TEST (TG)		WAVELENGTH (SUB/MAIN) (700)/(500)	
ASSAY CODE (1 POINT)(10)()		DILUTION (00301) (99)	
ASSAY POINT (31) (0) (0) (0)			
		< S.TYPE 1 >	< S.TYPE 2 >
S. VOLUME (REGULAR)		(3) (0) (0)	(1) (0) (0)
S. VOLUME (DECREASE)		(2) (0) (0)	(1) (0) (0)
S. VOLUME (INCREASE)		(5) (0) (0)	(1) (0) (0)
ABS. LIMIT		(0) () (INCREASE)	
PROZONE LIMIT		(0) () (HIGHER)	
REAGENT	R1	(300) (0) (*) (0)	
	R2	(0) (0) (*) (0)	
	R3	(0) (0) (*) (0)	
	R4	(0) (0) (*) (0)	
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()	
AUTO CALIB.			
TIME OUT BLANK	(0)	STD LIMIT	(0.1)
SPAN	(0)	DUPLICATE LIMITE	(400)
2 POINT	(0)	SENSITIVITY LIMIT	(0)
FULL	(0)	S1 ABS LIMIT	(-32000) (32000)
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()
BOTTLE	(CANCEL)		

* : Enter position or code number. *** : Enter calibrator or standard value.
 \$\$\$: Enter expected values.

☞ Modification from the previous version


(06/2009)
AA-H911-TGMLN-3



TRIGLYCERIDES

Ref.: TRIG-0200 12 x 20 mL
 TRIG-0400 9 x 50 mL

**APPLICATION HITACHI 911
 PROPOSAL**


Instruction use : 

Working temperature : 37°C

TEST (TG) (*)		TEST NAME (TG)		UNITS (g/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (TRIGLYCERIDES)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
AGE M F									
(0) (Y) (0) (0) (0) (0)				(0) (0)					
(0) (Y) (0) (0) (0) (0)									
		(\$\$) (\$\$) (\$\$) (\$\$)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE		<S.TYPE 2>					
(0)-(10)				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (TG)		WAVELENGTH (SUB/MAIN) (700)/(546)							
ASSAY CODE (1 POINT)(10)()		DILUTION (00301) (99)							
ASSAY POINT (31) (0) (0) (0)									
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(2) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(5) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(0) () (HIGHER)							
REAGENT R1		(300) (0) (*) (0)							
R2		(0) (0) (*) (0)							
R3		(0) (0) (*) (0)							
R4		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(400)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(-32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.
 \$\$\$: Enter expected values.

*** : Enter calibrator or standard value.

 Modification from the previous version

(06/2009)
 AA-H911-TRIG-4



UREA UV SL

ONE REAGENT PROCEDURE

Ref. : URSL-0400	2 x 62.5 mL
URSL-0420	4 x 62.5 mL
URSL-0500	5 x 125 mL
☞ URSL-0407	2 x 62.5 mL + STD
☞ URSL-0427	4 x 62.5 mL + STD
☞ URSL-0507	5 x 125 mL + STD

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (UREA) (*)	TEST NAME (UREA)	UNITS (g/l)	
DATA MODE (1: ON BOARD)	REPORT NAME (UREA SL ONE)		
CONTROL INTERVAL (300)	INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)		
EXPECTED VALUE <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>	
AGE M F			
(0) (Y) (0) (0) (0) (0)			(0) (0)
(0) (Y) (0) (0) (0) (0)			
(§§) (§§) (§§) (§§)			
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>	
(0)-(3)		(0)-(0)	
STD (1)	CONC (0)	POS (*)	S. VOL (3)
(2)	(***)	(*)	(3)
(3-6)	(0)	(0)	(0)
PRE. DIL. (0)	VOL (0)	CODE (*)	LOT (1)
(0)	(0)	(*)	(2)
(0)	(0)	(*)	(3-6)
			QUALITATIVE (0)
			(0)
			(NO) ()
			()

TEST (UREA)	WAVELENGTH (SUB/MAIN) (415)/(340)
ASSAY CODE (2 POINT CIN)(4)()	DILUTION (00301) (99)
ASSAY POINT (4) (8) (0) (0)	
	< S.TYPE 1 >
S. VOLUME (REGULAR)	(3) (0) (0)
S. VOLUME (DECREASE)	(2) (0) (0)
S. VOLUME (INCREASE)	(6) (0) (0)
ABS. LIMIT	(5000) () (DECREASE)
PROZONE LIMIT	(0) () (LOWER)
REAGENT R1	(300) (0) (*) (0)
R2	(0) (0) (*) (0)
R3	(0) (0) (*) (0)
R4	(0) (0) (*) (0)
CALIBRATION TYPE	(LINEAR) (2) (2) (0) ()
AUTO CALIB.	
TIME OUT BLANK (0)	STD LIMIT (0.1)
SPAN (0)	DUPLICATE LIMITE (200)
2 POINT (0)	SENSITIVITY LIMIT (0)
FULL (0)	S1 ABS LIMIT (-32000) (32000)
CHANGE LOT (CANCEL)	COMPENSATED LIMIT ()
BOTTLE (CANCEL)	

* : Enter position or code number.
 §§§ : Enter expected values.

*** : Enter calibrator or standard value.

Modification from the previous version



UREA UV SL

TWO REAGENT PROCEDURE

Ref. : URSL-0400	2 x 62.5 mL
URSL-0420	4 x 62.5 mL
URSL-0500	5 x 125 mL
☞ URSL-0407	2 x 62.5 mL + STD
☞ URSL-0427	4 x 62.5 mL + STD
☞ URSL-0507	5 x 125 mL + STD

APPLICATION HITACHI 911 PROPOSAL

Instruction use :

Working temperature : 37°C

TEST (UREA) (*)	TEST NAME (UREA)	UNITS (g/l)	
DATA MODE (1: ON BOARD)	REPORT NAME (UREA SL TWO)		
CONTROL INTERVAL (300)	INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)		
EXPECTED VALUE <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>		
AGE M F			
(0) (Y) (0) (0) (0) (0)			(0) (0)
(0) (Y) (0) (0) (0) (0)			
(§§) (§§)	(§§) (§§)		
TECHNICAL LIMIT <S.TYPE 1>	EXPECTED VALUE <S.TYPE 2>		
(0)-(3)	(0)-(0)		
STD (1)	CONC (0)	POS (*)	S. VOL (3)
(2)	(**)	(*)	(3)
(3-6)	(0)	(0)	(0)
			PRE. DIL. (0)
			VOL (0)
			CODE (*)
			LOT (1)
			QUALITATIVE (0)
			(NO) ()

TEST (UREA)	WAVELENGTH (SUB/MAIN) (415)/(340)
ASSAY CODE (2 POINT CIN)(3) ()	DILUTION (00301) (99)
ASSAY POINT (9) (13) (0) (0)	
	< S.TYPE 1 >
S. VOLUME (REGULAR)	< S.TYPE 2 >
(3) (0) (0)	(1) (0) (0)
S. VOLUME (DECREASE)	(1) (0) (0)
(2) (0) (0)	(1) (0) (0)
S. VOLUME (INCREASE)	(1) (0) (0)
(6) (0) (0)	(1) (0) (0)
ABS. LIMIT	(5000) () (DECREASE)
(0) () (LOWER)	
PROZONE LIMIT	
(0) () ()	
REAGENT R1	(240) (0) (*) (0)
R2	(60) (0) (*) (0)
R3	(0) (0) (*) (0)
R4	(0) (0) (*) (0)
CALIBRATION TYPE	(LINEAR) (2) (2) (0) ()
AUTO CALIB.	
TIME OUT BLANK (0)	STD LIMIT (0.1)
SPAN (0)	DUPLICATE LIMITE (200)
2 POINT (0)	SENSITIVITY LIMIT (0)
FULL (0)	S1 ABS LIMIT (-32000) (32000)
CHANGE LOT (CANCEL)	COMPENSATED LIMIT ()
BOTTLE (CANCEL)	

* : Enter position or code number.

*** : Enter calibrator or standard value.

§§§ : Enter expected values.



Modification from the previous version



UREA UV

Ref. : URUV-0400 9 x 50 mL
URUV-0500 6 x 100 mL

APPLICATION HITACHI 911 PROPOSAL

Instruction use : 
Working temperature : 37°C

TEST	(UREA) (*)	TEST NAME	(UREA)	UNITS	(g/l)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(UREA)						
CONTROL INTERVAL	(300)	INSTRUM. FACTOR	(Y=aX+b)	a	(1.0) b (0)				
EXPECTED VALUE	<S.TYPE 1>	EXPECTED VALUE			<S.TYPE 2>				
	AGE M F								
	(0) (Y) (0) (0) (0) (0)				(0) (0)				
	(0) (Y) (0) (0) (§§) (§§)				(0) (0)				
TECHNICAL LIMIT	<S.TYPE 1>	EXPECTED VALUE			<S.TYPE 2>				
	(0)-(3)				(0)-(0)				
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(3)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(3)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(0)	(0)	(0)	(*)	(3-6)	(0)	()
TEST	(UREA)	WAVELENGTH	(SUB/MAIN)	(415)/(340)					
ASSAY CODE	(2 POINT CIN)(4)()	DILUTION	(00301)	(99)					
ASSAY POINT	(4) (8) (0) (0)								
		< S.TYPE 1 >			< S.TYPE 2 >				
S. VOLUME (REGULAR)		(3) (0) (0)			(1) (0) (0)				
S. VOLUME (DECREASE)		(2) (0) (0)			(1) (0) (0)				
S. VOLUME (INCREASE)		(6) (0) (0)			(1) (0) (0)				
ABS. LIMIT		(5000) () (DECREASE)							
PROZONE LIMIT		(0) () (LOWER)							
REAGENT	R1	(300) (0) (*) (0)							
	R2	(0) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT			(0.1)				
SPAN	(0)	DUPLICATE LIMITE			(200)				
2 POINT	(0)	SENSITIVITY LIMIT			(0)				
FULL	(0)	S1 ABS LIMIT			(-32000) (32000)				
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT			()				
BOTTLE	(CANCEL)								

* : Enter position or code number.
§§§ : Enter expected values.

*** : Enter calibrator or standard value.



URIC ACID MONO SL

Ref. : AUML-0420	6 x 50 mL
AUML-0500	6 x 100 mL
AUML-0700	4 x 250 mL
☞ AUML-0427	6 x 50 mL + STD
☞ AUML-0507	6 x 100 mL + STD
☞ AUML-0707	4 x 250 mL + STD

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (AUML) (*)		TEST NAME (URAC)		UNITS (mg/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (URIC ACID)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE	M	F							
(0) (Y)	(0) (0)	(0) (0)	(0) (0)						
(0) (Y)	(0) (0)	(0) (0)							
	(§§) (§§)	(§§) (§§)							
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(250)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (URAC)			WAVELENGTH (SUB/MAIN) (700)/(546)						
ASSAY CODE (1 POINT)(10)()			DILUTION (00301) (99)						
ASSAY POINT (31) (0) (0) (0)									
			< S.TYPE 1 >	< S.TYPE 2 >					
S. VOLUME (REGULAR)			(7) (0) (0)	(1) (0) (0)					
S. VOLUME (DECREASE)			(3) (0) (0)	(1) (0) (0)					
S. VOLUME (INCREASE)			(14) (0) (0)	(1) (0) (0)					
ABS. LIMIT			(0) () (INCREASE)						
PROZONE LIMIT			(32000) (32000) (HIGHER)						
REAGENT R1			(300) (0) (*) (0)						
R2			(0) (0) (*) (0)						
R3			(0) (0) (*) (0)						
R4			(0) (0) (*) (0)						
CALIBRATION TYPE			(LINEAR) (2) (2) (0) ()						
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT		(0.1)				
SPAN (0)			DUPLICATE LIMITE		(200)				
2 POINT (0)			SENSITIVITY LIMIT		(0)				
FULL (0)			S1 ABS LIMIT		(- 32000) (32000)				
CHANGE LOT (CANCEL)			COMPENSATED LIMIT		()				
BOTTLE (CANCEL)									

* : Enter position or code number.
 §§§ : Enter expected values.

*** : Enter calibrator or standard value.

Modification from the previous version

(06/2009)
AA-H911-AUML-2



URIC ACID SL

Ref. : AUSL-0400 2 x 62.5 mL
 AUSL-0420 4 x 62.5 mL
 AUSL-0600 5 x 125 mL

ONE REAGENT PROCEDURE

APPLICATION HITACHI 911 PROPOSAL

Instruction use :
Working temperature : 37°C

TEST (URAC) (*)		TEST NAME (URAC)		UNITS (mg/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (URIC ACID)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE M		F							
(0) (Y) (0) (0)		(0) (0)		(0) (0)					
(0) (Y) (0) (0)		(0) (0)							
		(\$\$) (\$\$)		(\$\$) (\$\$)					
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
(0)-(250)		(0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(**)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (URAC)				WAVELENGTH (SUB/MAIN) (700)/(546)					
ASSAY CODE (1 POINT)(10)()				DILUTION (00301) (99)					
ASSAY POINT (31) (0) (0) (0)									
S. VOLUME (REGULAR)			< S.TYPE 1 >			< S.TYPE 2 >			
(7) (0) (0)			(1) (0) (0)			(1) (0) (0)			
S. VOLUME (DECREASE)			(3) (0) (0)			(1) (0) (0)			
S. VOLUME (INCREASE)			(14) (0) (0)			(1) (0) (0)			
ABS. LIMIT			(0) () (INCREASE)						
PROZONE LIMIT			(32000) (32000) (HIGHER)						
REAGENT									
R1			(300) (0) (*) (0)						
R2			(0) (0) (*) (0)						
R3			(0) (0) (*) (0)						
R4			(0) (0) (*) (0)						
CALIBRATION TYPE				(LINEAR) (2) (2) (0) ()					
AUTO CALIB.									
TIME OUT BLANK (0)			STD LIMIT			(0.1)			
SPAN (0)			DUPLICATE LIMITE			(200)			
2 POINT (0)			SENSITIVITY LIMIT			(0)			
FULL (0)			S1 ABS LIMIT			(- 32000) (32000)			
CHANGE LOT (CANCEL)			COMPENSATED LIMIT			()			
BOTTLE (CANCEL)									

* : Enter position or code number. *** : Enter calibrator or standard value.
 \$\$\$: Enter expected values.



URIC ACID SL

Ref. : AUSL-0400 2 x 62.5 mL
 AUSL-0420 4 x 62.5 mL
 AUSL-0600 5 x 125 mL

TWO REAGENT PROCEDURE

**APPLICATION HITACHI 911
 PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (URAC) (*)		TEST NAME (URAC)		UNITS (mg/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (URIC ACID)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(250)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(6)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(6)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (URAC)		ASSAY CODE (2 POINT)(10)()		WAVELENGTH (SUB/MAIN) (700)/(546)					
		ASSAY POINT (15) (31) (0) (0)		DILUTION (00301) (99)					
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(6) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(10) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(32000) (32000) (HIGHER)							
REAGENT		R1							
		(240) (0) (*) (0)							
		R2							
		(0) (0) (*) (0)							
		R3							
		(60) (0) (*) (0)							
		R4							
		(0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK		(0)		STD LIMIT (0.1)					
SPAN		(0)		DUPLICATE LIMITE (200)					
2 POINT		(0)		SENSITIVITY LIMIT (0)					
FULL		(0)		S1 ABS LIMIT (- 32000) (32000)					
CHANGE LOT		(CANCEL)		COMPENSATED LIMIT ()					
BOTTLE		(CANCEL)							

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



URIC ACID

☞ Ref. : ACUR-0200 12 x 20 mL
ACUR-0400 9 x 50 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :

Working temperature : 37°C

TEST (URAC) (*)		TEST NAME (URAC)		UNITS (mg/l)					
DATA MODE (1: ON BOARD)		REPORT NAME (URIC ACID)							
CONTROL INTERVAL (300)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT <S.TYPE 1>		EXPECTED VALUE <S.TYPE 2>							
		(0)-(250)							
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(*)	(7)	(0)	(0)	(*)	(1)	(0)	()
(2)	(***)	(*)	(7)	(0)	(0)	(*)	(2)	(0)	()
(3-6)	(0)	(0)	(10)	(0)	(0)	(*)	(3-6)	(0)	()
TEST (URAC)		ASSAY CODE (1 POINT)(10)()		WAVELENGTH (SUB/MAIN) (700)/(505)					
		ASSAY POINT (31) (0) (0) (0)		DILUTION (00301) (99)					
		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (REGULAR)		(7) (0) (0)		(1) (0) (0)					
S. VOLUME (DECREASE)		(3) (0) (0)		(1) (0) (0)					
S. VOLUME (INCREASE)		(14) (0) (0)		(1) (0) (0)					
ABS. LIMIT		(0) () (INCREASE)							
PROZONE LIMIT		(32000) (32000) (HIGHER)							
REAGENT		R1 (300) (0) (*) (0)							
		R2 (0) (0) (*) (0)							
		R3 (0) (0) (*) (0)							
		R4 (0) (0) (*) (0)							
CALIBRATION TYPE		(LINEAR) (2) (2) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT		(0.1)					
SPAN (0)		DUPLICATE LIMITE		(200)					
2 POINT (0)		SENSITIVITY LIMIT		(0)					
FULL (0)		S1 ABS LIMIT		(- 32000) (32000)					
CHANGE LOT (CANCEL)		COMPENSATED LIMIT		()					
BOTTLE (CANCEL)									

* : Enter position or code number.

*** : Enter calibrator or standard value.

\$\$\$: Enter expected values.



Modification from the previous version

(06/2009)
AA-H911-ACUR-4




APPLICATION HITACHI 911

PROPOSAL

For more details, see the Instructions for use: 
 Working temperature: 37°C

PROCEDURE

Preparation of working reagent :

R1 : Reagent 1  Ref :IAPA-6125 Ready to use
 R3 : Reagent 2 Ref :IAPA-5025 Ready to use

Calibration :


APOA1/B IP CALIBRATOR H Ref : IAPO-0042

The analyser prepare automatically successive 1/2 dilution in NaCl 9 g/L solution.

Add a zero point (NaCl 9g/L) : POS STD (1)

TEST (IAPA) (*)		TEST NAME (APOA1)		UNITS (mg/dL)					
DATA MODE (1: ON BOARD)		REPORT NAME (APOA1)							
CONTROL INTERVAL (0)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <SERUM> (*)-(*)		EXPECTED VALUE <URINE>(*)-(*)							
TECHNICAL LIMIT <SERUM> (0)-(conc STD(6))		EXPECTED VALUE <URINE> (0)-(0)							
STD	CONC	POS	S. VOL	PRE. DIL.	DIL.	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(6)	(0)	(0)	(*)	(*)	(0)	()
(2)	(*)	(*)	(2)	(3)	(318)	(*)	(*)	(0)	()
(3)	(*)	(*)	(2)	(3)	(158)	(*)	(*)	(0)	()
(4)	(*)	(*)	(5)	(3)	(195)	(*)	(*)	(0)	()
(5)	(*)	(*)	(10)	(3)	(190)	(*)	(*)	(0)	()
(6)	(*)	(*)	(10)	(6)	(190)	(*)	(*)	(0)	()
TEST (IAPA)		WAVELENGTH (SUB/MAIN) (700)/(340)							
ASSAY CODE (2 POINT END)(10)()		DILUTION (*) (99)							
ASSAY POINT (15) (31) (0) (0)		< SERUM >		< URINE >					
S. VOLUME (REGULAR)		(5) (6) (95)							
S. VOLUME (DECREASE)		(5) (3) (95)							
S. VOLUME (INCREASE)		(0) (0) (0)							
ABS. LIMIT		(-32000) (32000) (INCREASE)							
PROZONE LIMIT		(-32000) (32000) (UPPER)							
REAGENT R1		(250) (0) (*) (0)							
R2		(0) (0) () (0)							
R3		(40) (0) (*) (0)							
R4		(0) (0) () (0)							
CALIBRATION TYPE		(SPLINE) (6)-(0) (0) ()							
AUTO CALIB.									
TIME OUT BLANK (0)		STD LIMIT (999)							
SPAN (0)		DUPLICATE LIMITE (2000)							
2 POINT (0)		SENSITIVITY LIMIT (0)							
FULL (0)		S1 ABS LIMIT (- 32000) (32000)							
CHANGE LOT (CANCEL)		COMPENSATED LIMIT ()							
BOTTLE (CANCEL)									

* : Enter by the user

 Important modification from the previous version

(06/2009)
 AA-H911-IAPA-3


APPLICATION HITACHI 911

PROPOSAL

For more details, see the Instructions for use: 
 Working temperature: 37°C

PROCEDURE

Preparation of working reagent :

R1 : Reagent 1  Ref :IAPB-6125 Ready to use
 R3 : Reagent 2 Ref :IAPB-5025 Ready to use

Calibration :


APOA1/B IP CALIBRATOR H Ref : IAPO-0042

The analyser prepare automatically successive 1/2 dilution in NaCl 9g/L solution.

Add a zero point (NaCl 9g/L) : POS STD (1)

TEST	(IAPB) (*)	TEST NAME	(APOB)	UNITS	(mg/dL)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(APOB)						
CONTROL INTERVAL	(0)	INSTRUM. FACTOR	(Y=aX+b) a (1.0) b (0)						
EXPECTED VALUE	<SERUM> (*)-(*)	EXPECTED VALUE	<URINE> (*)-(*)						
TECHNICAL LIMIT	<SERUM> (0)-(conc STD(6))	EXPECTED VALUE	<URINE> (0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	DIL.	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(8)	(0)	(0)	(*)	(*)	(0)	()
(2)	(*)	(*)	(5)	(2)	(195)	(*)	(*)	(0)	()
(3)	(*)	(*)	(5)	(2)	(95)	(*)	(*)	(0)	()
(4)	(*)	(*)	(10)	(2)	(90)	(*)	(*)	(0)	()
(5)	(*)	(*)	(10)	(4)	(90)	(*)	(*)	(0)	()
(6)	(*)	(*)	(10)	(8)	(90)	(*)	(*)	(0)	()
TEST (IAPB)		WAVELENGTH (SUB/MAIN)	(700)/(340)						
ASSAY CODE (2 POINT END)	(10)()	DILUTION (*) (99)							
ASSAY POINT (15) (31) (0) (0)		< SERUM >						< URINE >	
S. VOLUME (REGULAR)		(10) (8) (90)							
S. VOLUME (DECREASE)		(10) (4) (90)							
S. VOLUME (INCREASE)		(0) (0) (0)							
ABS. LIMIT		(-32000) (32000) (INCREASE)							
PROZONE LIMIT		(-32000) (32000) (UPPER)							
REAGENT	R1	(220) (0) (*) (0)							
	R2	(0) (0) () (0)							
	R3	(40) (0) (*) (0)							
	R4	(0) (0) () (0)							
CALIBRATION TYPE		(SPLINE) (6)-(0) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(999)						
SPAN	(0)	DUPLICATE LIMITE	(2000)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(- 32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter by the user

 Important modification from the previous version

(06/2009)
 AA-H911-IAPB-3

APPLICATION HITACHI 911

PROPOSAL

For more details, see the Instructions for use: 
 Working temperature: 37°C

PROCEDURE

Preparation of working reagent :

R1 : Reagent 1 Ref :ICRP-6125 Ready to use

R3 : Reagent 2 Ref :ICRP-5025 Ready to use

Calibration :

CRP IP Calibrator Set Ref : ICRP-0043 Ready to use

CRP IP Calibrator H Ref : ICRP-0042 Ready to use

(The analyser prepares automatically successive dilution 1/2 in NaCl 9g/L solution.)

In both cases, add a zero point (NaCl9g/L) : POS STD (1)

TEST	(CRP) (*)	TEST NAME	(CRP)	UNITS	(mg/dl)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(CRP)						
CONTROL INTERVAL	(0)	INSTRUM. FACTOR (Y=aX+b)	a (1.0) b (0)						
EXPECTED VALUE	<SERUM > (*)-(*)	EXPECTED VALUE	<URINE>(*)-(*)						
TECHNICAL LIMIT	<SERUM> (0)-(conc STD(6))	EXPECTED VALUE	<URINE> (0)-(0)						
<i>Calibration with CRP IP CALIBRATOR H</i>									
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(18)	(16)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal /16)	(26)	(10)	(8)	(70)	(*)	(*)	(0)	()
(3)	(Conc Cal /8)	(26)	(25)	(8)	(75)	(*)	(*)	(0)	()
(4)	(Conc Cal /4)	(26)	(25)	(8)	(25)	(*)	(*)	(0)	()
(5)	(Conc Cal /2)	(26)	(8)	(0)	(0)	(*)	(*)	(0)	()
(6)	(Conc.Cal)	(26)	(16)	(0)	(0)	(*)	(*)	(0)	()
<i>Calibration with CRP IP CALIBRATOR SET</i>									
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal1)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(3)	(Conc Cal2)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(4)	(Conc Cal3)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(5)	(Conc Cal4)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(6)	(Conc.Cal5)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
TEST (CRP)		WAVELENGTH (SUB/MAIN)	(700)/(340)						
ASSAY CODE (2 POINT END)	(10)()	DILUTION (*)	(99)						
ASSAY POINT (15) (31) () ()		< SERUM >						< URINE >	
S. VOLUME (REGULAR)		(16) (0) (0)							
S. VOLUME (DECREASE)		(4) (0) (0)							
S. VOLUME (INCREASE)		(0) (0) (0)							
ABS. LIMIT		(-32000) (32000) (INCREASE)							
PROZONE LIMIT		(-32000) (32000) (UPPER)							
REAGENT	R1	(220) (0) (*) (0)							
	R2	(0) (0) () (0)							
	R3	(25) (0) (*) (0)							
	R4	(0) (0) () (0)							
CALIBRATION TYPE		(SPLINE) (6)-(0) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(999)						
SPAN	(0)	DUPLICATE LIMITE	(2000)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(- 32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter by the user

§§§ : Enter expected values

(08/2007)

AA-H911-ICRP-1

APPLICATION HITACHI 911
PROPOSAL

For more details, see the Instructions for use: 
 Working temperature: 37°C

PROCEDURE
Preparation of working reagent :

R1 : Reagent 1 Ref : IHAP-6125 Ready to use
 R3 : Reagent 2 Ref : IHAP-5025 Ready to use

Calibration :

PROTEIN IP Calibrator Set Ref : IPRO-0043 Ready to use
PROTEIN IP CALIBRATOR H Ref : IPRO-0041 / IPRO-0042
 Use pure and diluted at 1/2, 1/4, 1/8, 1/16 in NaCl 9 g/L solution.
 In both cases, add a zero point (NaCl9g/L) : POS STD (1)

TEST (IHAP) (*)				TEST NAME (HAP)		UNITS (mg/dL)			
DATA MODE (1: ON BOARD)				REPORT NAME (Haptoglobin)					
CONTROL INTERVAL (0)				INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)					
EXPECTED VALUE <SERUM> (*)-(*)				EXPECTED VALUE <URINE>(*)-(*)					
TECHNICAL LIMIT <SERUM>				EXPECTED VALUE <URINE>					
(0)-(conc STD(6))				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(2)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(3)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(4)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(5)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(6)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()

TEST (IHAP)		WAVELENGTH (SUB/MAIN) (700)/(340)	
ASSAY CODE (2 POINT END)(10)()		DILUTION (*) (99)	
ASSAY POINT (15) (31) (0) (0)		< SERUM >	
S. VOLUME (REGULAR)		(2) (0) (0)	
S. VOLUME (DECREASE)		(1) (0) (0)	
S. VOLUME (INCREASE)		(0) (0) (0)	
ABS. LIMIT		(-32000) (32000) (INCREASE)	
PROZONE LIMIT		(-32000) (32000) (UPPER)	
REAGENT	R1	(280) (0) (*) (0)	
	R2	(0) (0) () (0)	
	R3	(40) (0) (*) (0)	
	R4	(0) (0) () (0)	
CALIBRATION TYPE		(SPLINE) (6)-(0) (0) ()	
AUTO CALIB.			
TIME OUT BLANK	(0)	STD LIMIT	(999)
SPAN	(0)	DUPLICATE LIMITE	(2000)
2 POINT	(0)	SENSITIVITY LIMIT	(0)
FULL	(0)	S1 ABS LIMIT	(- 32000) (32000)
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()
BOTTLE	(CANCEL)		

* : Enter by the user

HbA1c

MANUAL LYSE

Ref.: HBAC-0240 1 x 32 mL

**APPLICATION HITACHI 911
PROPOSAL**

Instruction use :
Working temperature : 37°C

Preparation of samples

Dispense 1 mL of reagent R3 in a plastic or glass tube. Add 20µL of well mixed blood sample (calibrator, control or patient sample). Allow to stand for 5 minutes or until lysis is complete.

TEST (HbA1c) (*)		TEST NAME (HbA1c)		UNITS (%)					
DATA MODE (1: ON BOARD)		REPORT NAME (haemoglobin A1c)							
CONTROL INTERVAL (1000)		INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)							
EXPECTED VALUE <S.TYPE 1>			EXPECTED VALUE <S.TYPE 2>						
AGE M		F							
(0) (Y) (0) (0)		(0) (0)		(0) (0)					
(0) (Y) (0) (0)		(0) (0)							
		(\$\$) (\$\$)		(\$\$) (\$\$)					
TECHNICAL LIMIT <SERUM>		EXPECTED VALUE		<S.TYPE 2>					
				(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(18)	(6)	(0)	(0)	(501)	(1)	(0)	()
(2)	(***)	(*)	(6)	(0)	(0)	(*)	(2)	(0)	()
(3)	(***)	(*)	(6)	(0)	(0)	(*)	(3)	(0)	()
(4)	(***)	(*)	(6)	(0)	(0)	(*)	(4)	(0)	()
(5)	(***)	(*)	(6)	(0)	(0)	(*)	(5)	(0)	()
(6)	(0)	(*)	(6)	(0)	(0)	(*)	(6)	(0)	()
TEST (HbA1c)									
ASSAY CODE (1 POINT)(10)(-)				WAVELENGTH (SUB/MAIN) (-)/(660)					
ASSAY POINT (31) (0) (0) (0)				DILUTION (00311) (0)					
			< S.TYPE 1 >			< S.TYPE 2 >			
S. VOLUME (REGULAR)			(6) (0) (0)			(\$\$) (0) (0)			
S. VOLUME (DECREASE)			(XX) (0) (0)			(\$\$) (0) (0)			
S. VOLUME (INCREASE)			(XX) (0) (0)			(\$\$) (0) (0)			
ABS. LIMIT			(0) () (INCREASE)						
PROZONE LIMIT			(0) () (LOWER)						
REAGENT			R1			(240) (0) (*) (0)			
			R2			(0) (0) (*) (0)			
			R3			(80) (0) (*) (0)			
			R4			(0) (0) (*) (0)			
CALIBRATION TYPE				(SPLINE) (5) (5) (0) ()					
AUTO CALIB.									
TIME OUT BLANK			(0)			STD LIMIT			(999)
SPAN			(0)			DUPLICATE LIMITE			(1000)
2 POINT			(0)			SENSITIVITY LIMIT			(0)
FULL			(0)			S1 ABS LIMIT			(- 32000) (32000)
CHANGE LOT			(-)			COMPENSATED LIMIT			()
BOTTLE			(BLANK)						

* : Enter position or code number. \$\$: Enter expected values.
 *** : Enter calibrator or standard value. XX : User defined

For optimal performance, cuvette washes are recommended with this assay. Using the parameter Job screen and the Special wash programming function, schedule a cell wash using a Cell Clean/1% Hitergent mixture as the wash solution. Request that 250 µL and 100 µL of solution be dispensed from the R1 and R2 reagent probes after the HbA1c assay.



HbA1c

ON-BOARD LYSE

APPLICATION HITACHI 911 PROPOSAL

Instruction use : 
Working temperature : 37°C

TEST	(HbA1c) (*)	TEST NAME	(HbA1c)	UNITS	(%)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(hemoglobin A1c)						
CONTROL INTERVAL	(1000)	INSTRUM. FACTOR	(Y=aX+b) a (1.0) b (0)						
EXPECTED VALUE	<S.TYPE 1>		EXPECTED VALUE	<S.TYPE 2>					
	AGE	M	F						
	(0) (Y)	(0) (0)	(0) (0)		(0) (0)				
	(0) (Y)	(0) (0)	(0) (0)						
		(\$\$) (\$\$)	(\$\$) (\$\$)						
TECHNICAL LIMIT	<SERUM>		EXPECTED VALUE	<S.TYPE 2>					
	(\$\$)-(\$\$)			(0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0)	(18)	(4)	(4)	(200)	(501)	(1)	(0)	()
(2)	(**)	(*)	(4)	(4)	(200)	(*)	(2)	(0)	()
(3)	(**)	(*)	(4)	(4)	(200)	(*)	(3)	(0)	()
(4)	(**)	(*)	(4)	(4)	(200)	(*)	(4)	(0)	()
(5)	(**)	(*)	(4)	(4)	(200)	(*)	(5)	(0)	()
(6)	(0)	(*)	(4)	(4)	(200)	(*)	(6)	(0)	()
TEST	(HbA1c)		WAVELENGTH (SUB/MAIN)	(-)/(660)					
ASSAY CODE	(1 POINT)(10)(-)		DILUTION	(0) (0)					
ASSAY POINT	(31) (0) (0) (0)								
S. VOLUME (REGULAR)		< S.TYPE 1 >		< S.TYPE 2 >					
S. VOLUME (DECREASE)		(4) (4) (200)		(\$\$) (0) (0)					
S. VOLUME (INCREASE)		(XX) (0) (0)		(\$\$) (0) (0)					
ABS. LIMIT		(XX) (0) (0)		(\$\$) (0) (0)					
PROZONE LIMIT		(0) () (INCREASE)							
REAGENT	R1	(0) () (LOWER)							
	R2	(190) (0) (*) (0)							
	R3	(0) (0) (*) (0)							
	R4	(63) (0) (*) (0)							
		(0) (0) (*) (0)							
CALIBRATION TYPE		(SPLINE) (5) (5) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT		(999)					
SPAN	(0)	DUPLICATE LIMITE		(1000)					
2 POINT	(0)	SENSITIVITY LIMIT		(0)					
FULL	(0)	S1 ABS LIMIT		(- 32000) (32000)					
CHANGE LOT	(-)	COMPENSATED LIMIT		()					
BOTTLE	(BLANK)								

* : Enter position or code number.
 *** : Enter HbA1c Calibrator value.
 \$\$\$: Enter expected values.
 XX : User defined.

For optimal performance, cuvette washes are recommended with this assay. Using the parameter Job screen and the Special wash programming function, schedule a cell wash using a Cell Clean/1% Hitergent mixture as the wash solution. Request that 250 µL and 100 µL of solution be dispensed from the R1 and R2 reagent probes after the HbA1c assay.



APPLICATION HITACHI 911

For more details, see the Instructions for use: 
Working temperature: 37°C

PROCEDURE**Preparation of working reagent :**

R1 : Reagent 1 Ref :IMAL-6125 Ready to use
R3 : Reagent 2 Ref :IMAL-5025 Ready to use

Calibration :

μALBUMIN IP CALIBRATOR SET Ref : IMAL-0043 Ready to use
μALBUMIN IP Calibrator H Ref : IMAL-0042 Ready to use
(The analyser prepare automatically successive dilution 1/2 in NaCl 9g/L solution.)

In both cases, add a zero point (NaCl9g/L) : POS STD (1)

PERFORMANCE DATA**- Analytical range**

The reagent is linear from 0.3 to 40.0 mg/dL.
The exact range depends on the value of the calibrator used.

- Hook effect

No risk of error linked to hook effect up to 200 mg/dL.

- Precision**Within-run reproducibility**

Human urine 1	n=20	m= 3.2 mg/dL	CV=0.9 %
Human urine 2	n=20	m= 9.2 mg/dL	CV=0.9 %
Human urine 3	n=20	m= 19.6 mg/dL	CV=1.5%

Between-run reproducibility

Human urine 1	n=13	m= 2.2 mg/dL	CV=2.9%
Human urine 2	n=13	m= 9.1 mg/dL	CV=0.7%
Human urine 3	n=13	m= 19.2 mg/dL	CV=0.5 %

- Correlation

A comparative study was performed between ELITECH μALBUMIN IP and another commercial reagent (nephelometric method) on human urine samples. The parameters of linear regression are as follows :

Correlation coefficient : (r) = 0.9978
Linear regression : y = 1.0096x – 0.02 mg/dL.

- Interferences

According to SFBC recommendations, studies were performed to determine the level of interference from different compounds :

Haemoglobin : No significant interference up to 1000 mg/dL (10 g/L).

Bilirubin : No significant interference up to 15 mg/dL (150 mg/L, 257 μmol/L).

Other compounds may interfere. ⁽¹⁻²⁾

BIBLIOGRAPHY

- 1-Young,D.S., *Effects of preanalytical variables on clinical laboratory tests*, 2^{ème} Ed., AACC Press, (1997).
- 2- Young,D.S., *Effects of drugs on clinical laboratory tests*, 4^{ème} Ed., AACC Press, (1995).

(09/2007)
AA-H911-IMAL-1 (1/2)



TEST	(IMAL) (*)	TEST NAME	(IMAL)	UNITS	(mg/dl)
DATA MODE	(1: ON BOARD)	REPORT NAME	(IMAL)		
CONTROL INTERVAL	(0)	INSTRUM. FACTOR	(Y=aX+b) a (1.0) b (0)		
EXPECTED VALUE	<SERUM> (*)-(*)	EXPECTED VALUE	<URINE> (*)-(*)		
TECHNICAL LIMIT	<SERUM> (0)-(0)	EXPECTED VALUE	<URINE> (0)-(conc STD(6))		

Calibration with μ ALBUMIN IP CALIBRATOR H

STD	CONC	POS	S. VOL	PRE. DIL.	DIL.	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal /16)	(*)	(20)	(10)	(180)	(*)	(*)	(0)	()
(3)	(Conc Cal /8)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(4)	(Conc Cal /4)	(*)	(4)	(0)	(0)	(*)	(*)	(0)	()
(5)	(Conc Cal /2)	(*)	(8)	(0)	(0)	(*)	(*)	(0)	()
(6)	(Conc.Cal)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()

Calibration with μ ALBUMIN IP CALIBRATOR SET

STD	CONC	POS	S. VOL	PRE. DIL.	DIL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal1)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(3)	(Conc Cal2)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(4)	(Conc Cal3)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(5)	(Conc Cal4)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()
(6)	(Conc.Cal5)	(*)	(16)	(0)	(0)	(*)	(*)	(0)	()

TEST (IMAL)

ASSAY CODE (2 POINT END)(10)()

WAVELENGTH (SUB/MAIN) (700)/(340)

ASSAY POINT (16) (31) () ()

DILUTION (*) (99)

< SERUM >

< URINE >

S. VOLUME (REGULAR)

(16) (0) (0)

S. VOLUME (DECREASE)

(8) (0) (0)

S. VOLUME (INCREASE)

(0) (0) (0)

ABS. LIMIT

(-32000) (32000) (INCREASE)

PROZONE LIMIT

(-32000) (32000) (UPPER)

REAGENT

R1

(250) (0) (*) (0)

R2

(0) (0) () (0)

R3

(40) (0) (*) (0)

R4

(0) (0) () (0)

CALIBRATION TYPE

(LOG-LOGIT (4P)) (6)-(0) (0) ()

AUTO CALIB.

TIME OUT BLANK

(0)

STD LIMIT

(999)

SPAN

(0)

DUPLICATE LIMITE

(2000)

2 POINT

(0)

SENSITIVITY LIMIT

(0)

FULL

(0)

S1 ABS LIMIT

(- 32000) (32000)

CHANGE LOT

(CANCEL)

COMPENSATED LIMIT

()

BOTTLE

(CANCEL)

* : Enter by the user



APPLICATION HITACHI 911

For more details, see the Instructions for use: 
Working temperature: 37°C

PROCEDURE**Preparation of working reagent :**

R1 : Reagent 1 Ref : IIGB-6125 Ready to use

R3 : Reagent 2 Ref : IIGA-5025 Ready to use

Calibration :

PROTEIN IP Calibrator Set Ref : IPRO-0043 Ready to use

PROTEIN IP CALIBRATOR H Ref : IPRO-0041 / IPRO-0042 Ready to use

(The analyser prepare automatically successive dilution 1/2 in NaCl 9g/L solution.)

In both cases, add a zero point (NaCl 9g/L) : POS STD (1)

PERFORMANCE DATA**- Analytical range**

The reagent is linear from 2 to 650 mg/dL.

The exact range depends on the value of the calibrator used.

- Hook effect

No risk of error linked to hook effect up to 6000 mg/dL.

- Precision**Within-run reproducibility**

Low level n=20 m= 104 mg/dL CV=0.8 %

Medium level n=20 m= 217 mg/dL CV=1.0 %

High level n=20 m= 646 mg/dL CV=1.2%

Between-run reproducibility

Low level n=20 m= 107 mg/dL CV=1.7%

Medium level n=20 m= 222 mg/dL CV=1.4%

High level n=20 m= 649 mg/dL CV=1.1%

- Correlation

A comparative study was performed between ELITECH IgA IP and another commercial reagent (nephelometric method) on human serum samples. The parameters of linear regression are as follows :

Correlation coefficient : (r) = 0.9964

Linear regression : y = 1.0317x + 3 mg/dL.

- Interferences

According to SFBC recommendations, studies were performed to determine the level of interference from different compounds :

Bilirubin : No significant interference up to 15 mg/dL (150 mg/L, 257 µmol/L).

Hemoglobin : No significant interference up to 1000 mg/dL (10 g/L).

Triglycerides : No significant interference up to 2500 mg/dL (25 g/L, 28.5 mmol/L).

Heparin : No significant interference up to 50 mg/dL (0.5 g/L).

Sodium citrate : No significant interference up to 1000 mg/dL (10 g/L).

EDTA : No significant interference up to 5 mg/dL (0.05 g/L).

Other compounds may interfere. (1-2)

BIBLIOGRAPHY

1-Young,D.S., Effects of preanalytical variables on clinical laboratory tests, 2nd Ed., AACC Press, (1997).

2- Young,D.S., Effects of drugs on clinical laboratory tests, 4th Ed., AACC Press, (1995).

(11/2007)

AA-H911-IIGA-1 (p1/2)



TEST	(IGA) (*)	TEST NAME	(IGA)	UNITS	(mg/dl)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(IgA)						
CONTROL INTERVAL	(0)	INSTRUM. FACTOR	(Y=aX+b) a (1.0) b (0)						
EXPECTED VALUE	<SERUM> (*)-(*)	EXPECTED VALUE	<URINE>(*)-(*)						
TECHNICAL LIMIT	<SERUM> (0)-(conc STD(6))	EXPECTED VALUE	<URINE> (0)-(0)						
<i>Calibration with</i> PROTEIN IP CALIBRATOR H									
STD	CONC	POS	S. VOL	PRE. DIL.	DIL.	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal /16)	(*)	(5)	(5)	(195)	(*)	(*)	(0)	()
(3)	(Conc Cal /8)	(*)	(5)	(10)	(195)	(*)	(*)	(0)	()
(4)	(Conc Cal /4)	(*)	(10)	(10)	(190)	(*)	(*)	(0)	()
(5)	(Conc Cal /2)	(*)	(20)	(10)	(180)	(*)	(*)	(0)	()
(6)	(Conc.Cal)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
<i>Calibration with</i> PROTEIN IP Calibrator Set									
STD	CONC	POS	S. VOL	PRE. DIL.	DIL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal1)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(3)	(Conc Cal2)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(4)	(Conc Cal3)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(5)	(Conc Cal4)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(6)	(Conc.Cal5)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
TEST (IGA)									
ASSAY CODE	(2 POINT END)(10)()	WAVELENGTH (SUB/MAIN)	(700)/(340)						
ASSAY POINT	(15) (31) (0) (0)	DILUTION (*)	(99)						
		< SERUM >		< URINE >					
S. VOLUME (REGULAR)		(2) (0) (0)							
S. VOLUME (DECREASE)		(1) (0) (0)							
S. VOLUME (INCREASE)		(0) (0) (0)							
ABS. LIMIT		(0) (32000) (INCREASE)							
PROZONE LIMIT		(-32000) (32000) (UPPER)							
REAGENT	R1	(350) (0) (*) (0)							
	R2	(0) (0) () (0)							
	R3	(60) (0) (*) (0)							
	R4	(0) (0) () (0)							
CALIBRATION TYPE		(SPLINE) (6)-(0) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(999)						
SPAN	(0)	DUPLICATE LIMITE	(2000)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(- 32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter by the user



APPLICATION HITACHI 911

For more details, see the Instructions for use: 
Working temperature: 37°C

PROCEDURE**Preparation of working reagent :**

R1 : Reagent 1 Ref : IIGB-6125 Ready to use

R3 : Reagent 2 Ref : IIGG-5025 Ready to use

Calibration :

PROTEIN IP Calibrator Set Ref : IPRO-0043 Ready to use

PROTEIN IP CALIBRATOR H Ref : IPRO-0041 / IPRO-0042 Ready to use

(The analyser prepare automatically successive dilution 1/2 in NaCl 9g/L solution.)

In both cases, add a zero point (NaCl 9g/L) : POS STD (1)

PERFORMANCE DATA**- Analytical range**

The reagent is linear from 100 to 2600 mg/dL.
The exact range depends on the value of the calibrator used.

- Hook effect

No risk of error linked to hook effect up to 25000 mg/dL.

- Precision**Within-run reproducibility**

Low level	n=20	m= 422 mg/dL	CV=1.5 %
Medium level	n=20	m= 854 mg/dL	CV=1.5 %
High level	n=20	m= 2584 mg/dL	CV=1.4 %

Between-run reproducibility

Low level	n=20	m= 433 mg/dL	CV=1.2 %
Medium level	n=20	m= 867 mg/dL	CV=1.3 %
High level	n=20	m= 2632 mg/dL	CV=1.6 %

- Correlation

A comparative study was performed between ELITECH IgG IP and another commercial reagent (nephelometric method) on human serum samples. The parameters of linear regression are as follows :

Correlation coefficient : $(r) = 0.9899$

Linear regression : $y = 0.9786x - 55 \text{ mg/dL}$.

- Interferences

According to SFBC recommendations, studies were performed to determine the level of interference from different compounds :

Bilirubin : No significant interference up to 15 mg/dL (150 mg/L, 257 $\mu\text{mol/L}$).

Hemoglobin : No significant interference up to 1000 mg/dL (10 g/L).

Triglycerides : No significant interference up to 2500 mg/dL (25 g/L, 28.5 mmol/L).

Heparin : No significant interference up to 50 mg/dL (0.5 g/L).

Sodium citrate : No significant interference up to 1000 mg/dL (10 g/L).

EDTA : No significant interference up to 5 mg/dL (0.05 g/L).

Other compounds may interfere. ⁽¹⁻²⁾

BIBLIOGRAPHY

1-Young,D.S., *Effects of preanalytical variables on clinical laboratory tests*, 2nd Ed., AACC Press, (1997).

2- Young,D.S., *Effects of drugs on clinical laboratory tests*, 4th Ed., AACC Press, (1995).

(11/2007)

AA-H911-IIGG-1 (p1/2)



TEST (IGG) (*)	TEST NAME (IGG)	UNITS (mg/dl)							
DATA MODE (1: ON BOARD)	REPORT NAME (IgG)								
CONTROL INTERVAL (0)	INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)								
EXPECTED VALUE <SERUM> (*)-(*)	EXPECTED VALUE <URINE>(*)-(*)								
TECHNICAL LIMIT <SERUM>	EXPECTED VALUE <URINE>								
(0)-(conc STD(6))	(0)-(0)								
<i>Calibration with</i> PROTEIN IP CALIBRATOR H									
STD	CONC	POS	S. VOL	PRE. DIL.	DIL.	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(5)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal /16)	(*)	(2)	(5)	(318)	(*)	(*)	(0)	()
(3)	(Conc Cal /8)	(*)	(2)	(5)	(158)	(*)	(*)	(0)	()
(4)	(Conc Cal /4)	(*)	(5)	(5)	(195)	(*)	(*)	(0)	()
(5)	(Conc Cal /2)	(*)	(10)	(5)	(190)	(*)	(*)	(0)	()
(6)	(Conc.Cal)	(*)	(20)	(5)	(180)	(*)	(*)	(0)	()
<i>Calibration with</i> PROTEIN IP Calibrator Set									
STD	CONC	POS	S. VOL	PRE. DIL.	DIL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(5)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal1)	(*)	(20)	(5)	(180)	(*)	(*)	(0)	()
(3)	(Conc Cal2)	(*)	(20)	(5)	(180)	(*)	(*)	(0)	()
(4)	(Conc Cal3)	(*)	(20)	(5)	(180)	(*)	(*)	(0)	()
(5)	(Conc Cal4)	(*)	(20)	(5)	(180)	(*)	(*)	(0)	()
(6)	(Conc.Cal5)	(*)	(20)	(5)	(180)	(*)	(*)	(0)	()
TEST (IGG)									
ASSAY CODE (2 POINT END)(10)()	WAVELENGTH (SUB/MAIN) (700)/(340)								
ASSAY POINT (15) (31) (0) (0)	DILUTION (*) (99)								
	< SERUM >	< URINE >							
S. VOLUME (REGULAR)	(20) (5) (180)								
S. VOLUME (DECREASE)	(10) (5) (190)								
S. VOLUME (INCREASE)	(20) (10) (180)								
ABS. LIMIT	(0) (32000) (INCREASE)								
PROZONE LIMIT	(-32000) (32000) (UPPER)								
REAGENT	R1	(250) (0) (*) (0)							
	R2	(0) (0) () (0)							
	R3	(25) (0) (*) (0)							
	R4	(0) (0) () (0)							
CALIBRATION TYPE	(SPLINE) (6)-(0) (0) ()								
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT (999)							
SPAN	(0)	DUPLICATE LIMITE (2000)							
2 POINT	(0)	SENSITIVITY LIMIT (0)							
FULL	(0)	S1 ABS LIMIT (- 32000) (32000)							
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT ()							
BOTTLE	(CANCEL)								

* : Enter by the user

(11/2007)
AA-H911-IIGG-1 (p2/2)



APPLICATION HITACHI 911

For more details, see the Instructions for use: 
Working temperature: 37°C

PROCEDURE**Preparation of working reagent :**

R1 : Reagent 1 Ref : IIGB-6125 Ready to use

R3 : Reagent 2 Ref : IIGM-5025 Ready to use

Calibration :

PROTEIN IP Calibrator Set Ref : IPRO-0043 Ready to use

PROTEIN IP CALIBRATOR H Ref : IPRO-0041 / IPRO-0042 Ready to use

(The analyser prepare automatically successive dilution 1/2 in NaCl 9g/L solution.)

In both cases, add a zero point (NaCl 9g/L) : POS STD (1)

PERFORMANCE DATA**- Analytical range**

The reagent is linear from 10 to 450 mg/dL.

The exact range depends on the value of the calibrator used.

- Hook effect

No risk of error linked to hook effect up to 4500 mg/dL.

- Precision**Within-run reproducibility**

Low level n=20 m= 72 mg/dL CV=1.4 %

Medium level n=20 m= 147 mg/dL CV=1.6 %

High level n=20 m= 436 mg/dL CV=0.6 %

Between-run reproducibility

Low level n=20 m= 67 mg/dL CV=3.2 %

Medium level n=20 m= 142 mg/dL CV=3.7 %

High level n=20 m= 434 mg/dL CV=2.5 %

- Correlation

A comparative study was performed between ELITECH IgM IP and another commercial reagent (nephelometric method) on human serum samples. The parameters of linear regression are as follows :

Correlation coefficient : (r) = 0.9966

Linear regression : y = 1.0059x + 9 mg/dL.

- Interferences

According to SFBC recommendations, studies were performed to determine the level of interference from different compounds :

Bilirubin : No significant interference up to 15 mg/dL (150 mg/L, 257 µmol/L).

Hemoglobin : No significant interference up to 1000 mg/dL (10 g/L).

Triglycerides : No significant interference up to 2500 mg/dL (25 g/L, 28.5 mmol/L).

Heparin : No significant interference up to 50 mg/dL (0.5 g/L).

Sodium citrate : No significant interference up to 1000 mg/dL (10 g/L).

EDTA : No significant interference up to 5 mg/dL (0.05 g/L).

Other compounds may interfere. ⁽¹⁻²⁾

BIBLIOGRAPHY

1-Young,D.S., Effects of preanalytical variables on clinical laboratory tests, 2nd Ed., AACC Press, (1997).

2- Young,D.S., Effects of drugs on clinical laboratory tests, 4th Ed., AACC Press, (1995).

(11/2007)

AA-H911-IIGM-1 (p1/2)



TEST	(IGM) (*)	TEST NAME	(IGM)	UNITS	(mg/dl)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(IgM)						
CONTROL INTERVAL	(0)	INSTRUM. FACTOR	(Y=aX+b)	a (1.0) b (0)					
EXPECTED VALUE	<SERUM> (*)-(*)	EXPECTED VALUE		<URINE>(*)-(*)					
TECHNICAL LIMIT	<SERUM> (0)-(conc STD(6))	EXPECTED VALUE		<URINE> (0)-(0)					
<i>Calibration with</i> PROTEIN IP CALIBRATOR H									
STD	CONC	POS	S. VOL	PRE. DIL.	DIL.	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal /16)	(*)	(2)	(10)	(158)	(*)	(*)	(0)	()
(3)	(Conc Cal /8)	(*)	(5)	(10)	(195)	(*)	(*)	(0)	()
(4)	(Conc Cal /4)	(*)	(10)	(10)	(190)	(*)	(*)	(0)	()
(5)	(Conc Cal /2)	(*)	(20)	(10)	(180)	(*)	(*)	(0)	()
(6)	(Conc.Cal)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
<i>Calibration with</i> PROTEIN IP Calibrator Set									
STD	CONC	POS	S. VOL	PRE. DIL.	DIL.	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(2)	(Conc Cal1)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(3)	(Conc Cal2)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(4)	(Conc Cal3)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(5)	(Conc Cal4)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(6)	(Conc.Cal5)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
TEST (IGM)		WAVELENGTH (SUB/MAIN)	(700)/(340)						
ASSAY CODE	(2 POINT END)(10)()	DILUTION	(*) (99)						
ASSAY POINT	(15) (31) (0) (0)	< SERUM >		< URINE >					
S. VOLUME (REGULAR)		(2) (0) (0)							
S. VOLUME (DECREASE)		(20) (10) (180)							
S. VOLUME (INCREASE)		(4) (0) (0)							
ABS. LIMIT		(0) (32000) (INCREASE)							
PROZONE LIMIT		(-32000) (32000) (UPPER)							
REAGENT	R1	(350) (0) (*) (0)							
	R2	(0) (0) () (0)							
	R3	(50) (0) (*) (0)							
	R4	(0) (0) () (0)							
CALIBRATION TYPE		(SPLINE) (6)-(0) (0) ()							
AUTO CALIB.									
TIME OUT BLANK	(0)	STD LIMIT	(999)						
SPAN	(0)	DUPLICATE LIMITE	(2000)						
2 POINT	(0)	SENSITIVITY LIMIT	(0)						
FULL	(0)	S1 ABS LIMIT	(- 32000) (32000)						
CHANGE LOT	(CANCEL)	COMPENSATED LIMIT	()						
BOTTLE	(CANCEL)								

* : Enter by the user



APPLICATION HITACHI 911
PROPOSAL

 For more details, see the Instructions for use: 
 Working temperature: 37°C

PROCEDURE
Preparation of working reagent :

R1 : Reagent 1 Ref : IORO-6125 Ready to use

R3 : Reagent 2 Ref : IORO-5025 Ready to use

Calibration :

PROTEIN IP Calibrator Set Ref : IPRO-0043 Ready to use

PROTEIN IP CALIBRATOR H Ref : IPRO-0041 / IPRO-0042

Use pure and diluted at 1/2, 1/4, 1/8, 1/16 in NaCl 9 g/L solution.

In both cases, add a zero point (NaCl 9g/L) : POS STD (1)

TEST (IORO) (*)				TEST NAME (ORO)				UNITS (mg/dL)	
DATA MODE (1: ON BOARD)				REPORT NAME (Orosomucoid)					
CONTROL INTERVAL (0)				INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)					
EXPECTED VALUE <SERUM> (*)-(*)				EXPECTED VALUE <URINE> (*)-(*)					
TECHNICAL LIMIT <SERUM> (0)-(conc STD(6))				EXPECTED VALUE <URINE> (0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(2)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(3)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(4)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(5)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(6)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()

TEST (IORO)		WAVELENGTH (SUB/MAIN) (700)/(340)	
ASSAY CODE (2 POINT END)(10)()		DILUTION (*) (99)	
ASSAY POINT (15) (31) (0) (0)		< SERUM > < URINE >	
S. VOLUME (REGULAR)		(2) (0) (0)	
S. VOLUME (DECREASE)		(1) (0) (0)	
S. VOLUME (INCREASE)		(0) (0) (0)	
ABS. LIMIT		(-32000) (32000) (INCREASE)	
PROZONE LIMIT		(-32000) (32000) (UPPER)	
REAGENT R1		(290) (0) (*) (0)	
R2		(0) (0) () (0)	
R3		(40) (0) (*) (0)	
R4		(0) (0) () (0)	
CALIBRATION TYPE		(SPLINE) (5)-(0) (0) ()	
AUTO CALIB.			
TIME OUT BLANK (0)		STD LIMIT (999)	
SPAN (0)		DUPLICATE LIMITE (2000)	
2 POINT (0)		SENSITIVITY LIMIT (0)	
FULL (0)		S1 ABS LIMIT (- 32000) (32000)	
CHANGE LOT (CANCEL)		COMPENSATED LIMIT ()	
BOTTLE (CANCEL)			

* : Enter by the user

APPLICATION HITACHI 911
PROPOSAL

 For more details, see the Instructions for use: 
 Working temperature: 37°C

PROCEDURE
Preparation of working reagent :

R1 : Reagent 1 Ref.:IPAL-6125 Ready to use
 R3 : Reagent 2 Ref.:IPAL-5025 Ready to use

Calibration :

PROTEIN IP Calibrator Set Ref : IPRO-0043 Ready to use
 PROTEIN IP CALIBRATOR H Ref : IPRO-0041 / IPRO-0042
 Use pure and diluted at 1/2,1/4,1/8,1/16 in NaCl 9 g/L solution.
 In both cases, add a zero point (NaCl9g/L) : POS STD (1)

TEST	(IPAL) (*)	TEST NAME	(PAL)	UNITS	(mg/dL)				
DATA MODE	(1: ON BOARD)	REPORT NAME	(Prealbumine)						
CONTROL INTERVAL	(0)	INSTRUM. FACTOR	(Y=aX+b) a (1.0) b (0)						
EXPECTED VALUE	<SERUM> (*)-(*)	EXPECTED VALUE	<URINE>(*)-(*)						
TECHNICAL LIMIT	<SERUM> (0)-(conc STD(6))	EXPECTED VALUE	<URINE> (0)-(0)						
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(2)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(3)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(4)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(5)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
(6)	(*)	(*)	(2)	(0)	(0)	(*)	(*)	(0)	()
TEST (IPAL)									
ASSAY CODE (2 POINT END)(10)()					WAVELENGTH (SUB/MAIN) (700)/(340)				
ASSAY POINT (15) (31) (0) (0)					DILUTION (*)(99)				
					< SERUM >		< URINE >		
S. VOLUME (REGULAR)					(2) (0) (0)				
S. VOLUME (DECREASE)					(1) (0) (0)				
S. VOLUME (INCREASE)					(0) (0) (0)				
ABS. LIMIT					(-32000) (32000) (INCREASE)				
PROZONE LIMIT					(-32000) (32000) (UPPER)				
REAGENT R1					(220) (0) (*) (0)				
R2					(0) (0) () (0)				
R3					(20) (0) (*) (0)				
R4					(0) (0) () (0)				
CALIBRATION TYPE					(SPLINE) (5)-(0) (0) ()				
AUTO CALIB.									
TIME OUT BLANK (0)					STD LIMIT (999)				
SPAN (0)					DUPLICATE LIMITE (2000)				
2 POINT (0)					SENSITIVITY LIMIT (0)				
FULL (0)					S1 ABS LIMIT (- 32000) (32000)				
CHANGE LOT (CANCEL)					COMPENSATED LIMIT ()				
BOTTLE (CANCEL)									

* : Enter by the user

APPLICATION HITACHI 911
PROPOSAL

 For more details, see the Instructions for use: 
 Working temperature: 37°C

PROCEDURE
Preparation of working reagent :

R1 : Reagent 1 Ref.:ITRF-6125 Ready to use
 R3 : Reagent 2 Ref.:ITRF-5025 Ready to use

Calibration :

PROTEIN IP Calibrator Set Ref : IPRO-0043 Ready to use
 PROTEIN IP CALIBRATOR H Ref : IPRO-0041 / IPRO-0042
 Use pure and diluted at 1/2,1/4,1/8,1/16 in NaCl 9 g/L solution.
 In both cases, add a zero point (NaCl9g/L) : POS STD (1)

TEST (TRF) (*)				TEST NAME (TRF)			UNITS (mg/dL)		
DATA MODE (1: ON BOARD)				REPORT NAME (TRF)					
CONTROL INTERVAL (0)				INSTRUM. FACTOR (Y=aX+b) a (1.0) b (0)					
EXPECTED VALUE <SERUM> (*)-(*)				EXPECTED VALUE <URINE>(*)-(*)					
TECHNICAL LIMIT <SERUM> (0)-(conc STD(6))				EXPECTED VALUE <URINE> (0)-(0)					
STD	CONC	POS	S. VOL	PRE. DIL.	VOL	CODE	LOT	QUALITATIVE	(NO)
(1)	(0.00)	(*)	(3)	(0)	(0)	(*)	(*)	(0)	()
(2)	(*)	(*)	(10)	(3)	(190)	(*)	(*)	(0)	()
(3)	(*)	(*)	(10)	(3)	(190)	(*)	(*)	(0)	()
(4)	(*)	(*)	(10)	(3)	(190)	(*)	(*)	(0)	()
(5)	(*)	(*)	(10)	(3)	(190)	(*)	(*)	(0)	()
(6)	(*)	(*)	(10)	(3)	(190)	(*)	(*)	(0)	()

TEST (TRF)		WAVELENGTH (SUB/MAIN) (700)/(340)	
ASSAY CODE (2 POINT END)(10)()		DILUTION (*) (99)	
ASSAY POINT (15) (31) () ()		< SERUM >	
S. VOLUME (REGULAR)		(10) (3) (190)	
S. VOLUME (DECREASE)		(5) (3) (195)	
S. VOLUME (INCREASE)		(20) (3) (180)	
ABS. LIMIT		(-32000) (32000) (INCREASE)	
PROZONE LIMIT		(-32000) (32000) (UPPER)	
REAGENT R1		(250) (0) (*) (0)	
R2		(0) (0) () (0)	
R3		(20) (0) (*) (0)	
R4		(0) (0) () (0)	
CALIBRATION TYPE		(SPLINE) (6)-(0) (0) ()	
AUTO CALIB.			
TIME OUT BLANK (0)		STD LIMIT (999)	
SPAN (0)		DUPLICATE LIMITE (2000)	
2 POINT (0)		SENSITIVITY LIMIT (0)	
FULL (0)		S1 ABS LIMIT (- 32000) (32000)	
CHANGE LOT (CANCEL)		COMPENSATED LIMIT ()	
BOTTLE (CANCEL)			

* : Enter by the user