

ALT/GPT 4+1 SL

ONE REAGENT PROCEDURE

REF : ALSL-XXXX

Instruction for use :
 Consult Instruction for Use

APPLICATION - MINDRAY BS-120 / BS-200 / BS-200E - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Working reagent preparation: Mix 4 volumes of reagent R1 with 1 volume of reagent R2.
Working Reagent Stability: 5 days at 20-25°C / 2 weeks at 2-8°C

BS120 / BS200 / BS200E

PARAMETERS			
Test No.	ALSL **	R1 R2	200 / 200 / 200
Full Name	GPT	Sample volume	20 / 20 / 20
Standard No.	1	R1 Blank	
Reac. Type	Kinetic	Mixed Rgt Blank	
Pri. Wavelength	340 nm / 340 nm / 340 nm	Linearity Range	*
Sec. Wavelength		Linearity Limit	*
Direction	Decrease	Substrate Limit	*
Reac Time	3 10	Factor	
Incuba. Time		Prozone check	*
Units	U/L	q1	Q2
Precision	0.1	q3	Q4
		PC	Abs
CALIBRATION (Cal + Rgt Blk)			
Rule	Linear one point / Linear two point / Linear two point		
Sensitivity	1		
Replicates	2		
Interval (days)	0		
Difference Limit			
SD			
Blank Response			
Error Limit			
Correlation coefficient	1		

*Enter by the user.