

ELITech Clinical Systems SAS

CK NAC SL ONE REAGENT PROCEDURE

REF : CKSL-XXXX

<u>Instruction for use</u>:

Consult Instruction for Use

APPLICATION - MINDRAY BS-120 / BS-200 / BS-200E - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Working reagent preparation: Mix 4 volumes of reagent R1 with 1 volume of reagent R2. Working Reagent Stability: 2 days at 20-25 $^{\circ}$ C / 2 weeks at 2-8 $^{\circ}$ C

BS120 / BS200 / BS200E

| PARAMETERS | | | |
|-----------------------------|--|-----------------|-----------------|
| Test | CKSL | R1 | 250 / 250 / 250 |
| No. | ** | R2 | |
| Full Name | CK | Sample volume | 10/10/10 |
| Standard No. | 1 | R1 Blank | |
| Reac. Type | Kinetic | Mixed Rgt Blank | |
| Pri. Wavelength | 340 nm / 340 nm / 340 nm | Linearity Range | * |
| Sec. Wavelength | | Linearity Limit | * |
| Direction | Increase | Substrate Limit | * |
| Reac Time | 8 10 | Factor | |
| Incuba. Time | | Prozone check | * |
| Units | U/L | q1 | Q2 |
| Precision | Integer | q3 | Q4 |
| | | PC | Abs |
| CALIBRATION (Cal + Rgt Blk) | | | |
| Rule | Linear one point / Linear two point / Linear two point | | |
| Sensitivity | 1 | | |
| Replicates | 2 | | |
| Interval (days) | 0 | | |
| Difference Limit | | | |
| SD | | | |
| Blank Response | | | |
| Error Limit | | | |
| Correlation coefficient | 1 | | |

*Enter by the user.