

Instruction for use :
 Consult Instruction for Use

APPLICATION - MINDRAY BS-120 / BS-200 / BS-200E - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

BS120 / BS200 / BS200E

PARAMETERS			
Test No.	HDLL **	R1 R2	300 / 300 / 240 100 / 100 / 80
Full Name	HDL	Sample volume	3 / 3 / 2.4
Standard No.	1	R1 Blank	
Reac. Type	Endpoint	Mixed Rgt Blank	
Pri. Wavelength	578 nm / 578 nm / 570 nm	Linearity Range	*
Sec. Wavelength		Linearity Limit	*
Direction	Increase	Substrate Limit	*
Reac Time	0 20	Factor	
Incuba. Time	19	Prozone check	*
Units	mg/dL	q1	Q2
Precision	Integer	q3	Q4
		PC	Abs
CALIBRATION (Cal + Rgt Blk)			
Rule	Linear one point / Linear two point / Linear two point		
Sensitivity	1		
Replicates	2		
Interval (days)	0		
Difference Limit			
SD			
Blank Response			
Error Limit			
Correlation coefficient	1		

*Enter by the user.