

Instruction for use :  
 Consult Instruction for Use

**APPLICATION - MINDRAY BS-120 / BS-200 / BS-200E - PROPOSAL**

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

**BS120 / BS200 / BS200E**

<u>PARAMETERS</u>			
Test	IFRT	R1	180
No.	**	R2	45
Full Name	FERR	Sample volume	30
Standard No.	6	R1 Blank	
Reac. Type	Fixed T	Mixed Rgt Blank	
Pri. Wavelength	546 nm	Linearity Range	*
Sec. Wavelength		Linearity Limit	*
Direction	Increase	Substrate Limit	*
Reac Time	2 15	Factor	
Incuba. Time		Prozone check	*
Units	ng/mL	q1	Q2
Precision	Interger	q3	Q4
		PC	Abs
<u>CALIBRATION (Cal + Rgt Blk)</u>			
Rule	Spline		
Sensitivity	1		
Replicates	2		
Interval (days)	0		
Difference Limit			
SD			
Blank Response			
Error Limit			
Correlation coefficient	1		

\*Enter by the user.