

Instruction for use :
 Consult Instruction for Use

APPLICATION - MINDRAY BS-120 / BS-200 / BS-200E - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

BS120 / BS200 / BS200E

<u>PARAMETERS</u>			
Test	DTIBC	R1	220 / 220 / 220
No.	**	R2	66 /66 /66
Full Name	TIBC	Sample volume	18 / 18 / 18
Standard No.	1	R1 Blank	
Reac. Type	Fixed Time	Mixed Rgt Blank	
Pri. Wavelength	630 nm / 630nm / 630 nm	Linearity Range	*
Sec. Wavelength		Linearity Limit	*
Direction	Increase	Substrate Limit	*
Reac Time	2 15	Factor	
Incuba. Time		Prozone check	*
Units	µg/L	q1	Q2
Precision	Interger	q3	Q4
		PC	Abs
<u>CALIBRATION (Cal + Rgt Blk)</u>			
Rule	Linear one point / Linear two point / Linear two point		
Sensitivity	1		
Replicates	2		
Interval (days)	0		
Difference Limit			
SD			
Blank Response			
Error Limit			
Correlation coefficient	1		
Intercept	-50		

*Enter by the user.