

Consult Instruction for Use

APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Define / Edite Chemestries

Chem	<input type="text" value="AMSL"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Serum"/>
Chemistry	<input type="text" value="AMYLASE"/>			Print name	<input type="text"/>
Reaction Type	<input type="text" value="Kinetic"/>	Reaction Direction	<input type="text" value="Positive"/>		
Pri Wave	<input type="text" value="412 nm"/>	Sec Wave	<input type="text"/>		
Unit	<input type="text" value="U/L"/>	Decimal	<input type="text" value="0"/>		
Blank time	<input type="text"/>	Incubation Time	<input type="text" value="0"/>		
		Reaction Time	<input type="text" value="5"/>	<input type="text" value="16"/>	
Sample Vol	Standard	Aspirated	Diluent	Reagent Vol	
	<input type="text" value="2"/> µL	<input type="text"/> µL	<input type="text"/> µL	R1	<input type="text" value="200"/> µL
Decreased	<input type="text"/> µL	<input type="text"/> µL	<input type="text"/> µL	R2	<input type="text"/> µL
Increased	<input type="text"/> µL	<input type="text"/> µL	<input type="text"/> µL		
<input type="checkbox"/> Sample Blank		<input type="checkbox"/> Auto rerun			

Define / Edite Chemestries

Chem	<input type="text" value="AMSL"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Serum"/>
Chemistry	<input type="text" value="AMYLASE"/>			Print name	<input type="text"/>

Linearity range (standard)	<input type="text" value="*"/>	<input type="text" value="*"/>	Linearity Limit	<input type="text" value="*"/>	
Linearity range (Decreased)	<input type="text" value="*"/>	<input type="text" value="*"/>	Substrate Depletion	<input type="text"/>	
Linearity range (Increased)	<input type="text" value="*"/>	<input type="text" value="*"/>	Mixed Blank Abs	<input type="text" value="*"/>	<input type="text"/>
R1 Blank Abs	<input type="text" value="*"/>	<input type="text" value="*"/>	On-board Stability	<input type="text" value="*"/> Jour	
Blank Response	<input type="text" value="*"/>	<input type="text" value="*"/>	Reagent Alarm Limit	<input type="text"/>	
Twin Chemistry	<input type="text"/>				
<input type="checkbox"/> Enzyme Linear Extension					

<input type="checkbox"/> Prozone Check	<input type="checkbox"/> Rtae Check				
Q1 <input type="text"/>	Q2 <input type="text"/>	Q3 <input type="text"/>	Q4 <input type="text"/>	PC <input type="text"/>	ABS <input type="text"/>

Calibration setup

Chem	<input type="text" value="AMSL"/>		
Calibration settings			
Math Model	<input type="text" value="Two-point linear"/>		
Factor	<input type="text"/>		
Replicates	<input type="text" value="2"/>		
Acceptance Limits			
Cal Time	<input type="text"/> Heure		
Slope Diff	<input type="text"/>	SD	<input type="text"/>
Sensitivity	<input type="text"/>	Reproductb.	<input type="text"/>
Deter Coeff	<input type="text"/>		
Auto calib.			
<input type="checkbox"/> Bottle Changed	<input type="checkbox"/> Lot Changed	<input type="checkbox"/> Cal Time	

Calibrator	Conc.	Pos	Lot No.
Water	0	W	
Cal1	xxx	xxx	

*Values entered by the operator