

Consult Instruction for Use REF : ASSL-XXXX

APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Define / Edite Chemestries	
Chem ASSL N°	Sample Type Serum
Chemistry AST	Print name
Reaction Type Kinetic	Reaction Direction Negative
Pri Wave 340 nm	Sec Wave
Unit U/L	Decimal 0.1
	Incubation Time 28
Blank time	Reaction Time 5 16
Sample Vol Aspirated	Diluent Reagent Vol
Standard 10 µL µL	μL R1 [160 μL
Decreased μL μL	
Increased μL μL	μι
Sample Blank Auto rerun	
Define / Edite Chemestries	
Chem ASSL N°	Sample Type Serum
Chemistry AST	Print name
Linearity range (standard) * *	Linearity Limit *
Linearity range (Decreased) *	Substrate Depletion
Linearity range (Increased) * *	Mixed Blank Abs *
R1 Blank Abs * *	On-borad Stability * Jour
Blank Response * *	Reagent Alarm Limit
Twin Chemistry	Enzyme Linear Extension
Prozone Check ① Rtae Check	
Q1 Q2 Q3 Q3	Q4 PC ABS
Calibration setup	
Chem ASSL	
Calibration settings	Calibrator Conc. Pos Lot No.
Math Model Two-point linear	Water 0 W Cal1 xxx xxx
Factor Replicates 2	
ractor replicates 2	
Acceptance Limits	
Cal Time Heure	
Slope Diff SD SD	
Sensitivity Reproductb.	
Deter Coeff	
Auto calib.	
☐Bottle Changed ☐Lot Changed ☐Cal Time	