



Consult Instruction for Use

ELITech Clinical Systems SAS

REF: BITO-XXXX

## APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Define / Edite Chemestries	
Cl. DITO	South To Comme
Chem BITO N°	Sample Type Serum
Chemistry TOTAL BILIRUBIN	Print name
Reaction Type Endpoint	Reaction Direction Positive
Pri Wave 546 nm	Sec Wave 700 nm
Unit mg/dL	Decimal 0.01
	Incubation Time 28
Blank time -3 -1	Reaction Time 39 41
Sample Vol Aspirated	Diluent Reagent Vol
Standard 10 μL μL	μL R1 160 μL
Decreased μL μL	μL R2 40 μL
Increased µL µL	μL
Sample Blank Auto rerun	
Define / Edite Chemestries	
Chem BITO N°	Sample Type Serum
Chemistry TOTAL BILIRUBIN	Print name
Linearity range (standard) * *	Linearity Limit *
Linearity range (Decreased) * *	Substrate Depletion
Linearity range (Increased) *	Mixed Blank Abs *
R1 Blank Abs * *	On-borad Stability * Jour
Blank Response * *	Reagent Alarm Limit
Twin Chemistry	Enzyme Linear Extension
Prozone Check (0) Rtae Check	
Q1 Q2 Q3 Q3	Q4 PC ABS ABS
O. Phantha and a	
Calibration setup	
Chem BITO	
Calibration settings	Calibrator Conc. Pos Lot No.
Math Model Two-point linear	Water 0 W   Cal1 xxx xxx
Factor Replicates 2	
Acceptance Limits	
Cal Time Heure	
Slope Diff SD	
Sensitivity Reproductb.	
Deter Coeff	
Deter com	
Auto calib.	
□Bottle Changed □Cal Time	