

REF: CALA-XXXX



ELITech Clinical Systems SAS

Consult Instruction for Use

## APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Define / Edite Chemestries	
g Gu	
Chem CALA N°	Sample Type Serum
Chemistry CALCIUM	Print name
Reaction Type Endpoint	Reaction Direction Positive
Pri Wave 660 nm	Sec Wave 700 nm
Unit mg/dL	Decimal 0.01
	Incubation Time 0
Blank time [-3 -1	Reaction Time 27 29
Sample Vol Aspirated	Diluent Reagent Vol
Standard 2	μL R1 200 μL
Decreased μL μL	μL R2 μL
Increased µL µL	μ
Sample Blank Auto rerun	
Define / Edite Chemestries	
Chem CALA N°	Sample Type Serum
Chemistry CALCIUM	Print name
Linearity range (standard) *	Linearity Limit *
Linearity range (Decreased) * *	Substrate Depletion
Linearity range (Increased) *	Mixed Blank Abs *
R1 Blank Abs * *	On-borad Stability * Jour
Blank Response * *	Reagent Alarm Limit
Twin Chemistry	Enzyme Linear Extension
Prozone Check (0) Rtae Check	
Q1 Q2 Q3 Q3	Q4 PC ABS ABS
Calibration setup	
Chem CALA	
Calibration settings	Calibrator Conc. Pos Lot No.
Math Model Two-point linear	Water 0 W   Cal1 xxx xxx
	COIL AAA AAA
Factor Replicates 2	
Acceptance Limits	
Cal Time Heure	
Slope Diff SD	
Sensitivity Reproductb.	
Deter Coeff	
Auto calib.	
Dottle Changed Diet Changed DC-17:	
□Bottle Changed □Cal Time	