



ELITech Clinical Systems SAS REF : CHSL-XXXX

## Consult Instruction for Use

## APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Define / Edite Chemestries
Chem CHSL N° Sample Type Serum  Chemistry CHOLESTEROL Print name
Reaction Type Endpoint Reaction Direction Positive
Pri Wave 505 nm Sec Wave
Unit mg/dL Decimal 0
Incubation Time 0
Blank time -3 -1 Reaction Time 27 29
Sample Vol Aspirated Diluent Reagent Vol
Standard 2 μL μL R1 200 μL
Decreased μL μL μL R2 μL
Increased µL µL µL
Sample Blank Auto rerun
Define / Edite Chemestries
Chem CHSL N° Sample Type Serum
Chemistry CHOLESTEROL Print name
Linearity range (standard) *
Linearity range (Decreased) *
Linearity range (Increased) *
R1 Blank Abs *
Blank Response * * Reagent Alarm Limit
Twin Chemistry Enzyme Linear Extension
Twin chemistry
Prozone Check (0) Rtae Check
Q1 Q2 Q3 Q4 PC ABS
Calibration setup
Chem
Calibration settings Calibrator Conc. Pos Lot No.
Math Model Two-point linear Water 0 W Call xxx xxx
Factor Replicates 2
ractui
Acceptance Limits
Cal Time Heure
Slope Diff SD
Sensitivity Reproductb.
Deter Coeff
Auto calib.
□Bottle Changed □Lot Changed □Cal Time