

Consult Instruction for Use

REF : GISL-XXXX

APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Define / Edite Chemestries

Chem	<input type="text" value="GISL"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Serum"/>
Chemistry	<input type="text" value="GAMMA GT"/>	Print name	<input type="text"/>		
Reaction Type	<input type="text" value="Kinetic"/>	Reaction Direction	<input type="text" value="Positive"/>		
Pri Wave	<input type="text" value="412 nm"/>	Sec Wave	<input type="text"/>		
Unit	<input type="text" value="U/L"/>	Decimal	<input type="text" value="0"/>		
Blank time		<input type="text"/>	<input type="text"/>	Incubation Time	<input type="text" value="28"/>
				Reaction Time	<input type="text" value="5"/> <input type="text" value="16"/>
Sample Vol		Aspirated		Diluent	
Standard	<input type="text" value="10.9"/> μL	<input type="text"/> μL	<input type="text"/> μL	Reagent Vol	
Decreased	<input type="text"/> μL	<input type="text"/> μL	<input type="text"/> μL	R1	<input type="text" value="120"/> μL
Increased	<input type="text"/> μL	<input type="text"/> μL	<input type="text"/> μL	R2	<input type="text" value="30"/> μL
<input type="checkbox"/> Sample Blank		<input type="checkbox"/> Auto rerun			

Define / Edite Chemestries

Chem	<input type="text" value="GISL"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Serum"/>
Chemistry	<input type="text" value="GAMMA GT"/>	Print name	<input type="text"/>		

Linearity range (standard)	* <input type="text"/>	* <input type="text"/>	Linearity Limit	* <input type="text"/>
Linearity range (Decreased)	* <input type="text"/>	* <input type="text"/>	Substrate Depletion	<input type="text"/>
Linearity range (Increased)	* <input type="text"/>	* <input type="text"/>	Mixed Blank Abs	* <input type="text"/> <input type="text"/>
R1 Blank Abs	* <input type="text"/>	* <input type="text"/>	On-board Stability	* <input type="text"/> Jour
Blank Response	* <input type="text"/>	* <input type="text"/>	Reagent Alarm Limit	<input type="text"/>
Twin Chemistry	<input type="text"/>	<input type="checkbox"/> Enzyme Linear Extension		

<input type="checkbox"/> Prozone Check	<input checked="" type="radio"/> Rtae Check				
Q1 <input type="text"/>	Q2 <input type="text"/>	Q3 <input type="text"/>	Q4 <input type="text"/>	PC <input type="text"/>	ABS <input type="text"/>

Calibration setup

Chem	<input type="text" value="GISL"/>
------	-----------------------------------

Calibration settings			
Math Model	<input type="text" value="Two-point linear"/>		
Factor	<input type="text"/>	Replicates	<input type="text" value="2"/>

Acceptance Limits			
Cal Time	<input type="text"/>	Heure	
Slope Diff	<input type="text"/>	SD	<input type="text"/>
Sensitivity	<input type="text"/>	Reproductb.	<input type="text"/>
Deter Coeff	<input type="text"/>		

Auto calib.			
<input type="checkbox"/> Bottle Changed	<input type="checkbox"/> Lot Changed	<input type="checkbox"/> Cal Time	

Calibrator	Conc.	Pos	Lot No.
Water	0	W	
Cal1	xxx	xxx	

*Values entered by the operator