

Consult Instruction for Use

### APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

#### Define / Edite Chemestries

Chem	<input type="text" value="PROB"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Serum"/>
Chemistry	<input type="text" value="TOTAL PROTEIN"/>			Print name	<input type="text"/>
Reaction Type	<input type="text" value="Endpoint"/>	Reaction Direction	<input type="text" value="Positive"/>		
Pri Wave	<input type="text" value="546 nm"/>	Sec Wave	<input type="text"/>		
Unit	<input type="text" value="g/dL"/>	Decimal	<input type="text" value="0.01"/>		
Blank time	<input type="text" value="-3"/>	<input type="text" value="-1"/>	Incubation Time	<input type="text" value="0"/>	
			Reaction Time	<input type="text" value="60"/>	<input type="text" value="62"/>
Sample Vol	Aspirated	Diluent	Reagent Vol		
Standard	<input type="text" value="2"/> µL	<input type="text"/>	R1	<input type="text" value="200"/> µL	
Decreased	<input type="text"/>	<input type="text"/>	R2	<input type="text"/>	
Increased	<input type="text"/>	<input type="text"/>			
	<input type="checkbox"/> Sample Blank	<input type="checkbox"/> Auto rerun			

#### Define / Edite Chemestries

Chem	<input type="text" value="PROB"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Serum"/>
Chemistry	<input type="text" value="TOTAL PROTEIN"/>			Print name	<input type="text"/>

Linearity range (standard)	<input type="text" value="*"/>	<input type="text" value="*"/>	Linearity Limit	<input type="text" value="*"/>	
Linearity range (Decreased)	<input type="text" value="*"/>	<input type="text" value="*"/>	Substrate Depletion	<input type="text"/>	
Linearity range (Increased)	<input type="text" value="*"/>	<input type="text" value="*"/>	Mixed Blank Abs	<input type="text" value="*"/> <input type="text"/>	
R1 Blank Abs	<input type="text" value="*"/>	<input type="text" value="*"/>	On-board Stability	<input type="text" value="*"/> Jour	
Blank Response	<input type="text" value="*"/>	<input type="text" value="*"/>	Reagent Alarm Limit	<input type="text"/>	
Twin Chemistry	<input type="text"/>		<input type="checkbox"/> Enzyme Linear Extension		

<input type="checkbox"/> Prozone Check	<input type="checkbox"/> Rtae Check	<input type="checkbox"/> Q1	<input type="checkbox"/> Q2	<input type="checkbox"/> Q3	<input type="checkbox"/> Q4	<input type="checkbox"/> PC	<input type="checkbox"/> ABS
--	-------------------------------------	-----------------------------	-----------------------------	-----------------------------	-----------------------------	-----------------------------	------------------------------

#### Calibration setup

Chem	<input type="text" value="PROB"/>		
------	-----------------------------------	--	--

**Calibration settings**

Math Model

Factor  Replicates

**Acceptance Limits**

Cal Time  Heures

Slope Diff  SD

Sensitivity  Reproductb.

Deter Coeff

**Auto calib.**

Bottle Changed  Lot Changed  Cal Time

Calibrator	Conc.	Pos	Lot No.
Water	0	W	
Cal1	xxx	xxx	

\*Values entered by the operator