

Consult Instruction for Use

APPLICATION - Mindray - BS 360 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Define / Edite Chemestries

Chem	<input type="text" value="PRTU"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Urine"/>
Chemistry	<input type="text" value="PROTEIN U"/>			Print name	<input type="text"/>
Reaction Type	<input type="text" value="Endpoint"/>	Reaction Direction	<input type="text" value="Positive"/>		
Pri Wave	<input type="text" value="660 nm"/>	Sec Wave	<input type="text"/>		
Unit	<input type="text" value="mg/dL"/>	Decimal	<input type="text" value="0.1"/>		
Blank time	<input type="text" value="-3"/>	<input type="text" value="-1"/>	Incubation Time	<input type="text" value="0"/>	
			Reaction Time	<input type="text" value="27"/>	<input type="text" value="29"/>
Sample Vol	<input type="text"/>	Aspirated	<input type="text"/>	Diluent	<input type="text"/>
Standard	<input type="text" value="2"/> μ L	<input type="text"/>	<input type="text"/>	Reagent Vol	<input type="text"/>
Decreased	<input type="text"/>	<input type="text"/>	<input type="text"/>	R1	<input type="text" value="100"/> μ L
Increased	<input type="text"/>	<input type="text"/>	<input type="text"/>	R2	<input type="text"/>
<input type="checkbox"/> Sample Blank		<input type="checkbox"/> Auto rerun			

Define / Edite Chemestries

Chem	<input type="text" value="PRTU"/>	N°	<input type="text"/>	Sample Type	<input type="text" value="Urine"/>
Chemistry	<input type="text" value="PROTEIN U"/>			Print name	<input type="text"/>

Linearity range (standard)	<input type="text" value="*"/>	<input type="text" value="*"/>	Linearity Limit	<input type="text" value="*"/>	
Linearity range (Decreased)	<input type="text" value="*"/>	<input type="text" value="*"/>	Substrate Depletion	<input type="text"/>	
Linearity range (Increased)	<input type="text" value="*"/>	<input type="text" value="*"/>	Mixed Blank Abs	<input type="text" value="*"/>	<input type="text"/>
R1 Blank Abs	<input type="text" value="*"/>	<input type="text" value="*"/>	On-board Stability	<input type="text" value="*"/>	Jour
Blank Response	<input type="text" value="*"/>	<input type="text" value="*"/>	Reagent Alarm Limit	<input type="text"/>	
Twin Chemistry	<input type="text"/>				

<input type="checkbox"/> Prozone Check	<input type="checkbox"/> Rtae Check				
Q1 <input type="text"/>	Q2 <input type="text"/>	Q3 <input type="text"/>	Q4 <input type="text"/>	PC <input type="text"/>	ABS <input type="text"/>

Calibration setup

Chem	<input type="text" value="PRTU"/>
------	-----------------------------------

Calibration settings

Math Model

Factor

Replicates

Acceptance Limits

Cal Time Heure

Slope Diff SD

Sensitivity Reproductb.

Deter Coeff

Auto calib.

Bottle Changed Lot Changed Cal Time

Calibrator	Conc.	Pos	Lot No.
Water	0	W	
Cal1	xxx	xxx	

*Values entered by the operator