


CHOLESTEROL SL

Réf.: CHSL-XXXX

APPLICATION PRESTIGE 24i - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use

Instruction for use : 

Working temperature : 37°C

Item Name		XXX	CHSL	
DATA INFORMATION				
UNITS	mg/dL			
DECIMALS	0			
ANALYSIS				
TYPE	END			
Main W Length 1	505			
Sub W Length 1				
METHOD				
CORR	SLOPE	INTER		
Y=	1	X+	0	
CALIBRATION				
TYPE	Linear			
#1	xxx	#4		
#2		#5		
#3		#6		
NORMAL RANGE				
	MALE		FEMALE	
	Low	High	Low	High
Serum	xxx	xxx	xxx	xxx
Urine				
Plasma				
CSF				
Dialysis				
other				

Item Name		XXX	CHSL
ASPIRATION			
KIND	<input type="radio"/> single <input type="radio"/> double		
SAMPLE	VOLUME	μL	
3	300		
REAGENT 1 vol			
DATA PROCESS			
READ	Start	End	
Main	52	54	
Sub			
FACTOR			
Blank correction	1		
ABSORBANCE LIMIT			
Low	-3.000		
High	2.000		
Endpoint limit			
2			
Linear check (%)			
0			
Dilution			
Diluent	<input type="radio"/> 99:DI 1 <input type="radio"/> 100:DI 2		
PROZONE CHECK			
	start	End	Limit %
First			low
Second			<input type="radio"/>
third			<input type="radio"/>
			high
			<input type="radio"/>

Item name		XXX	CHSL
Auto Rerun SW			
<input type="radio"/> ON <input type="radio"/> OFF			
Auto Rerun Range (result)			
<input type="radio"/> ON <input type="radio"/> OFF			
Auto rerun condition (Absorbance)			
ABS. range		lower	<input type="radio"/> ON <input type="radio"/> OFF
		high	<input type="radio"/> ON <input type="radio"/> OFF
Prozone range		<input type="radio"/> ON <input type="radio"/> OFF	
	lower	higher	
Serum	20	600	
Urine			
Plasma			
CSF			
Dialysis			
other			

XXX: enter data by the user