

APPLICATION - RESPONS 920 - ERBA XL200 / XL-180 - EM200 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Consult Instruction for Use

TEST DETAILS

Test	:	ALP	Auto Rerun	<input type="checkbox"/>
Report Name	:	ALP IFCC	Online calibration	<input type="checkbox"/>
Unit	:	U/L	Cuvette Wash	<input type="checkbox"/>
Wavelength-Primary	:	405	Total reagents	: 2
Assay Type	:	RATE-A	Curve Type	: Linear
M1 Start	:	0	M1 End	: 0
M2 Start	:	22	M2 End	: 30
Sample Replicates	:	1	Standard replicates	: 3
Control Replicates	:	1	Control Interval	: 0
Reaction Direction	:	Increasing	React. Abs. Limit	: *
Prozone Limit %	:	0	Prozone Check	: lower
Linearity Limit %	:	0	Delta Abs./Min.	: 0.0000
Technical Minimum	:	*	Technical Maximum	: *
Y= aX + b	a=	1.0000	B=	0.0000

TEST VOLUMES

Test	:	ALP
Sample Type	:	Serum
Sample Volumes		
Normal	:	4.00 µL
Increase	:	* µL
Decrease	:	* µL
Standard Volume	:	4.00 µL
reagent Volumes and Stirrer Speed		
RGT-1 Volume	:	160 µL
RGT-2 Volume	:	40 µL
R1 Stirrer Speed	:	medium
R2 Stirrer Speed	:	medium

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

REFERENCE RANGES

Test	:	ALP
Sample Type	:	Serum
Reference Range	:	DEFAULT
Category	:	*
Reference Range		
	Lower Limit	Lower Limit
	U/L	U/L
Normal	:	*
Panic	:	*

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

*: Values entered by the operator