

APPLICATION - RESPONS 920 - ERBA XL200 / XL-180 - EM200 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Consult Instruction for Use

TEST DETAILS

Test	: CK	Auto Rerun	<input type="checkbox"/>
Report Name	: CK NAC SL	Online calibration	<input type="checkbox"/>
Unit	: U/L	Cuvette Wash	<input type="checkbox"/>
Wavelength-Primary	: 340	Total reagents	: 2
Assay Type	: RATE-A	Curve Type	: Linear
M1 Start	: 0	M1 End	: 0
M2 Start	: 25	M2 End	: 31
Sample Replicates	: 1	Standard replicates	: 3
Control Replicates	: 1	Control Interval	: 0
Reaction Direction	: Increasing	React. Abs. Limit	: *
Prozone Limit %	: 0	Prozone Check	: lower
Linearity Limit %	: 0	Delta Abs./Min.	: 0.0000
Technical Minimum	: *	Technical Maximum	: *
Y= aX + b	a= : 1.0000	B= :	0.0000

TEST VOLUMES

Test	: CK
Sample Type	: Serum
Sample Volumes	
Normal	: 8.00 µL
Increase	: * µL
Decrease	: * µL
Standard Volume	: 8.00 µL
reagent Volumes and Stirrer Speed	
RGT-1 Volume	: 160 µL
RGT-2 Volume	: 40 µL
R1 Stirrer Speed	: medium
R2 Stirrer Speed	: medium

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

REFERENCE RANGES

Test	: CK	
Sample Type	: Serum	
Reference Range	: DEFAULT	
Category	: *	
Reference Range		
	Lower Limit	Lower Limit
	U/L	U/L
Normal	: *	: *
Panic	: *	: *

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

*: Values entered by the operator