

## APPLICATION - RESPONS 920 - ERBA XL200 / XL-180 - EM200 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Consult Instruction for Use

### TEST DETAILS

Test	: Gluc	Auto Rerun	<input type="checkbox"/>
Report Name	: GLUCOSE PAP SL	Online calibration	<input type="checkbox"/>
Unit	: mg/dL	Decimal Places	: 1
Wavelength-Primary	: 505	Secondary	: -
Assay Type	: 1-Point	Curve Type	: Linear
M1 Start	: 0	M1 End	: 0
M2 Start	: 35*	M2 End	: 35*
Sample Replicates	: 1	Standard replicates	: 3
Control Replicates	: 1	Control Interval	: 0
Reaction Direction	: Increasing	React. Abs. Limit	: *
Prozone Limit %	: 0	Prozone Check	: lower
Linearity Limit %	: 0	Delta Abs./Min.	: 0.0000
Technical Minimum	: *	Technical Maximum	: *
Y= aX + b	a= : 1.0000	B= :	0.0000
		Reagent 1	: GPSL R1
		Reagent 2	:

### TEST VOLUMES

Test	: Gluc
Sample Type	: Serum
<b>Sample Volumes</b>	
Normal	: 2.00 µL
Increase	: * µL
Decrease	: * µL
Standard Volume	: 2.00 µL
<b>reagent Volumes and Stirrer Speed</b>	
RGT-1 Volume	: 200 µL
RGT-2 Volume	: 0 µL
R1 Stirrer Speed	: high
R2 Stirrer Speed	: 0

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

### REFERENCE RANGES

Test	: Gluc	
Sample Type	: Serum	
Reference Range	: DEFAULT	
Category	: *	
<b>Reference Range</b>		
	Lower Limit	Lower Limit
	mg/dL	mg/dL
Normal	: *	: *
Panic	: *	: *

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

\* XL-180 : 34

\*: Values entered by the operator