

APPLICATION - RESPONS 920 - ERBA XL200 / XL-180 - EM200 - PROPOSAL

This application is intended to serve as a guide for using the referenced ELITechGroup Reagent on this instrument system only. It is recommended that the user validate this application prior to routine use.

Consult Instruction for Use

TEST DETAILS

Test	:	PHO	Auto Rerun	<input type="checkbox"/>
Report Name	:	PHOSPHORUS	Online calibration	<input type="checkbox"/>
Unit	:	mg/dL	Cuvette Wash	<input type="checkbox"/>
Wavelength-Primary	:	340	Total reagents	: 1
Assay Type	:	1-Point	Curve Type	: Linear
M1 Start	:	0	M1 End	: 0
M2 Start	:	17	M2 End	: 17
Sample Replicates	:	1	Standard replicates	: 3
Control Replicates	:	1	Control Interval	: 0
Reaction Direction	:	Increasing	React. Abs. Limit	: *
Prozone Limit %	:	0	Prozone Check	: lower
Linearity Limit %	:	0	Delta Abs./Min.	: 0.0000
Technical Minimum	:	*	Technical Maximum	: *
Y= aX + b	a=	: 1.0000	B=	: 0.0000

TEST VOLUMES

Test	:	PHO
Sample Type	:	Serum
Sample Volumes		
Normal	:	2.00 µL
Increase	:	* µL
Decrease	:	* µL
Standard Volume	:	2.00 µL
reagent Volumes and Stirrer Speed		
RGT-1 Volume	:	200 µL
RGT-2 Volume	:	0 µL
R1 Stirrer Speed	:	medium
R2 Stirrer Speed	:	0

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

REFERENCE RANGES

Test	:	PHO
Sample Type	:	Serum
Reference Range	:	DEFAULT
Category	:	*
Reference Range		
	Lower Limit	Lower Limit
	mg/dL	mg/dL
Normal	:	*
Panic	:	*

Sample Types	
<input checked="" type="checkbox"/>	Serum
<input type="checkbox"/>	Urine
<input type="checkbox"/>	CSF
<input checked="" type="checkbox"/>	Plasma
<input type="checkbox"/>	Whole blood
<input type="checkbox"/>	Other

*: Values entered by the operator